



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

TSY  
2-4-2009

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1/9/09 Cal 1 0.103 AC Cal 2 0.102 AC


Between the dates of 12/12/08 and 1/9/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —  
The instrument was returned to service on —

Sarah Cione

Sarah L. Cione

1/12/09




ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 12<sup>th</sup>, day of January, 2009.

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

B4  
2.4.2009

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department  
DATE 1/9/09 TIME 8:49  
INTOXILYZER SERIAL # 80-002345 LOCATION Team 3-ODM

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( ✓ ) 1. Clock time check.
- ( ✓ ) 2. Date check.

OPERATIONAL TESTS

- ( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
- ( ✓ ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( ✓ ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( ✓ ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 102 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Sarah Cione

2 of 7<sub>sc</sub>

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
01/09/2009

08:51:10

154

Standard Lot#: 735401G  
Last Changed By: C.- CONLEY #34517

QAS: S. CIONE #50616  
TUCSON PD

Operator: S. CIONE #50616  
TUCSON PD

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:51:56
Diagnostic Test	Pass	08:52:23
Air Blank	0.000	08:52:51
0.100 Cal Check	0.102	08:53:09
Air Blank	0.000	08:53:39
Subject Test	0.000	08:54:00
Air Blank	0.000	08:54:27
Five Minute Wait	ABT*	
Air Blank	0.000	08:55:03

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

3 of 7 SC

154

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
01/09/2009 08:55:12

Standard Lot#: 735401G  
Last Changed By: C.- CONLEY #34517

QAS: S. CIONE #50616  
TUCSON PD

Operator: S. CIONE #50616  
TUCSON PD

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:55:37
Air Blank	IPS**	08:55:47

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

4 of 7 sc

154

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
01/09/2009

08:55:52

---

Standard Lot#: 735401G  
Last Changed By: C. CONLEY #34517

---

QAS: S. CIONE #50616  
TUCSON PD

---

Operator: S. CIONE #50616  
TUCSON PD

---

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	08:56:33
Diagnostic Test	Pass	08:57:01
Air Blank	0.000	08:57:28
0.100 Cal Check	0.102	08:57:47
Air Blank	0.000	08:58:16
Subject Test	0.047*	09:01:25
Air Blank	0.000	09:01:56
Air Blank	0.000	09:02:23
Subject Test	XXX**	09:02:41
Air Blank	0.000	09:03:14

\*Deficient Sample  
 \*\*Invalid Sample  
 Begin new deprivation period

**Not a Successfully Completed Test Sequence**

5 of 7 sc

TS4

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number: 80-002345  
01/09/2009 09:03:34

-----  
Standard Lot#: 735401G  
Last Changed By: C. CONLEY #34517  
-----

QAS: S. CIONE #50616  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	09:04:01
Cal Check	0.102	09:04:20
Air Blank	0.000	09:04:49

*Sarah Cione*

Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

6 of 7 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

BS4  
2.4.2009

QA Specialist Sarah Cione AGENCY Tucson Police Department  
DATE 1/9/09 TIME 8:49  
Intoxilyzer Serial # 80-002345 Location Team 3 - ODM

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C + 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 103 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number: 80-002345  
01/09/2009

08:49:37

7 of 7 SC

BH

-----  
Standard Lot#: 735401G  
Last Changed By: C. CONLEY #34517  
-----

-----  
QAS: S. CIONE #50616  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	08:50:04
Cal Check	0.103	08:50:23
Air Blank	0.000	08:50:52

*Sarah Cione*

Operator's Signature