



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

134  
5.11.2009

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5/1/09 Cal 1 0.103 AC Cal 2 0.103 AC

Between the dates of 4/2/09 and 5/1/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.  
The instrument was returned to service on —.

Sarah Cione

Sarah L. Cione


5/4/09

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 4, day of May, 20 09.

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

1 of 7sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

T54  
5.11.2009

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 5/1/09 TIME 10:10

INTOXILYZER SERIAL # 80-002345 LOCATION Team 3

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( ✓ ) 1. Clock time check.
- ( ✓ ) 2. Date check.

OPERATIONAL TESTS

- ( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
- ( ✓ ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( ✓ ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( ✓ ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 103 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

---



---



---



---

SIGNATURE Sarah Cione

2 of 7 sc

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
05/01/2009

10:12:28

SM

Standard Lot#: 826001I  
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616  
TUCSON PD

Operator: S. CIONE #50616  
TUCSON PD

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:13:11
Diagnostic Test	Pass	10:13:38
Air Blank	0.000	10:14:05
0.100 Cal Check	0.103	10:14:24
Air Blank	0.000	10:14:53
Subject Test	0.000	10:16:22
Air Blank	0.000	10:16:49
Five Minute Wait	ABT*	
Air Blank	0.000	10:17:24

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

3 of 7sc  
ASy

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
05/01/2009 10:17:29

---

Standard Lot#: 826001I  
Last Changed By: C. CONLEY #34517

---

QAS: S. CIONE #50616  
TUCSON PD

---

Operator: S. CIONE #50616  
TUCSON PD

---

Subject: SQAP,TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	10:17:58
Air Blank	IPS**	10:18:07

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1827, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7sc

104

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
05/01/2009 10:18:13

Standard Lot#: 826001I  
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616  
TUCSON PD

Operator: S. CIONE #50616  
TUCSON PD

Subject: SQAP,TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:18:53
Diagnostic Test	Pass	10:19:21
Air Blank	0.000	10:19:48
0.100 Cal Check	0.103	10:20:07
Air Blank	0.000	10:20:36
Subject Test	0.067*	10:23:45
Air Blank	0.000	10:24:16
Air Blank	0.000	10:24:44
Subject Test	XXX**	10:25:02
Air Blank	0.000	10:25:35

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

5 of 7sc

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number: 80-002345  
05/01/2009 10:25:50

184

-----  
Standard Lot#: 826001I  
Last Changed By: C. CONLEY #34517  
-----

QAS: S. CIONE #50616  
TUCSON PD -  
-----

Test	g/210L	Time
Air Blank	0.000	10:26:17
Cal Check	0.103	10:26:36
Air Blank	0.000	10:27:05

*Sarah Cione*

Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

6 of 7 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

JSy  
5.11.09

QA Specialist Sarah Cione AGENCY Tucson Police Department  
DATE 5/1/09 TIME 10:10  
Intoxilyzer Serial # 80-002345 Location Team 3

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 103 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7 SC

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number: 80-002345  
05/01/2009 10:10:58

TSU

-----  
Standard Lot#: 826001I  
Last Changed By: C. CONLEY #34517  
-----

QAS: S. CIONE #50616  
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:11:26
Cal Check	0.103	10:11:44
Air Blank	0.000	10:12:14

Sarah Cione  
Operator's Signature