



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSY
11-30-09

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/19/09 Cal 1 0.103 AC Cal 2 0.103 AC

Between the dates of 10/23/09 and 11/19/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub

11/25/09

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 25, day of November, 20 09

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

197
CS

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

By
11-30-09

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department
DATE 11/19/09 TIME 07:29
INTOXILYZER SERIAL # 80-002345 LOCATION TEAM 3

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 103 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Cheryl Straub

Pursuant to section 38-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

277
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
11/19/2009 07:32:36

AM

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:33:33 |
| Diagnostic Test | Pass | 07:34:01 |
| Air Blank | 0.000 | 07:34:28 |
| 0.100 Cal Check | 0.103 | 07:34:47 |
| Air Blank | 0.000 | 07:35:17 |
| Subject Test | 0.000 | 07:35:35 |
| Air Blank | ABT* | 07:35:41 |
| Air Blank | 0.000 | 07:36:10 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

377
CS

INTOXILYZER 8000
 Location: ARIZONA
 Serial Number : 80-002345
 11/19/2009 07:36:22

Standard Lot#: 826001I
 Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
 TUCSON PD

Operator: C. STRAUB #50772
 TPD

Subject: SQAP, SQAP
 DOB: 11/11/1111
 Sex: F Weight: 111

15 Minute Deprivation Period? Yes

BM

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 07:37:21 |
| Air Blank | IPS** | 07:37:30 |

*RFI Detect
 **Improper Sample

Not a Successfully Completed Test Sequence

487
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
11/19/2009

07:37:56

SH

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:38:50 |
| Diagnostic Test | Pass | 07:39:18 |
| Air Blank | 0.000 | 07:39:45 |
| 0.100 Cal Check | 0.103 | 07:40:04 |
| Air Blank | 0.000 | 07:40:33 |
| Subject Test | 0.124* | 07:43:42 |
| Air Blank | 0.000 | 07:44:14 |
| Air Blank | 0.000 | 07:44:41 |
| Subject Test | XXX** | 07:45:19 |
| Air Blank | 0.000 | 07:45:53 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

587
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
11/19/2009 07:46:13

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
TUCSON PD

TS4

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 07:46:40 |
| Cal Check | 0.103 | 07:46:59 |
| Air Blank | 0.000 | 07:47:28 |



Operator's Signature

EXHIBIT G-2

687
CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TS4
11-30-09

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
DATE 11/19/09 TIME 07:29
Intoxilyzer Serial # 80-002345 Location TEAM 3

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 103 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

787
CS

AS4

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
11/19/2009 07:29:56

Standard Lot#: 8260011
Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 07:30:24 |
| Cal Check | 0.103 | 07:30:42 |
| Air Blank | 0.000 | 07:31:12 |



Operator's Signature