



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 270 S. Stone Avenue Tucson, AZ 85701-1917

TSY
2-16-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 02/11/10 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 01/13/10 and 02/11/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
 The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub

02/11/10

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 11th, day of February, 2010.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

187
CS

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

TS4
2-16-2010

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department
DATE 02/11/10 TIME 07:01
INTOXILYZER SERIAL # 80-002345 LOCATION TEAM 3

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Cheryl Straub

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

297
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
02/11/2010

07:03:26

CS

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:04:18 |
| Diagnostic Test | Pass | 07:04:46 |
| Air Blank | 0.000 | 07:05:13 |
| 0.100 Cal Check | 0.101 | 07:05:32 |
| Air Blank | 0.000 | 07:06:01 |
| Subject Test | 0.000 | 07:06:19 |
| Air Blank | ABT* | 07:06:26 |
| Air Blank | 0.000 | 07:06:54 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

387
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
02/11/2010

07:06:58

ph

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 07:07:30 |
| Air Blank | IPS** | 07:07:39 |

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

497
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
02/11/2010

07:07:46

PH

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:08:40 |
| Diagnostic Test | Pass | 07:09:07 |
| Air Blank | 0.000 | 07:09:35 |
| 0.100 Cal Check | 0.102 | 07:09:54 |
| Air Blank | 0.000 | 07:10:23 |
| Subject Test | 0.185* | 07:13:32 |
| Air Blank | 0.000 | 07:14:04 |
| Air Blank | 0.000 | 07:14:32 |
| Subject Test | XXX** | 07:15:01 |
| Air Blank | 0.000 | 07:15:35 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

577
CS


INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
02/11/2010 07:16:01

ph

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 07:16:29 |
| Cal Check | 0.099 | 07:16:47 |
| Air Blank | 0.000 | 07:17:17 |



Operator's Signature

EXHIBIT G-2

687
CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

TS4
2-16-2010

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
DATE 02/11/10 TIME 07:01
Intoxilyzer Serial # 80-002345 Location TEAM 3

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

717
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
02/11/2010


07:01:56

Standard Lot#: 642845
Last Changed By: C. STRAUB #50772

ph

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 07:02:23 |
| Cal Check | 0.100 | 07:02:42 |
| Air Blank | 0.000 | 07:03:12 |


Operator's Signature