



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
4.18.2011

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 04/07/11 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 03/09/2011 and 04/07/2011 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub


04/18/2011

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 18th day of April, 2011.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

187 CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

T34
4.18.2011

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 04/07/2011 TIME 10:23

INTOXILYZER SERIAL # 80-002345 LOCATION CRIME LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Cheryl Straub

287
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
04/07/2011

10:25:05

134

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:25:53 |
| Diagnostic Test | Pass | 10:26:21 |
| Air Blank | 0.000 | 10:26:48 |
| 0.100 Cal Check | 0.101 | 10:27:07 |
| Air Blank | 0.000 | 10:27:36 |
| Subject Test | 0.000 | 10:28:05 |
| Air Blank | ABT* | 10:28:13 |
| Air Blank | 0.000 | 10:28:41 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

377
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
04/07/2011 10:28:48

BH

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 10:29:17 |
| Air Blank | IPS** | 10:29:26 |

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

487
CS

134

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
04/07/2011

10:29:34

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:30:20 |
| Diagnostic Test | Pass | 10:30:47 |
| Air Blank | 0.000 | 10:31:14 |
| 0.100 Cal Check | 0.101 | 10:31:33 |
| Air Blank | 0.000 | 10:32:03 |
| Subject Test | 0.181* | 10:35:11 |
| Air Blank | 0.000 | 10:35:44 |
| Air Blank | 0.000 | 10:36:11 |
| Subject Test | XXX** | 10:36:25 |
| Air Blank | 0.000 | 10:36:59 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

587
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
04/07/2011 10:38:55

TS4

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 10:39:22 |
| Cal Check | 0.100 | 10:39:41 |
| Air Blank | 0.000 | 10:40:11 |



Operator's Signature

697 CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

BY
4.18.2011

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
DATE 04/07/2011 TIME 10:23
Intoxilyzer Serial # 80-002345 Location CRIME LAB

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clyde Sta

DPS Form Exh G-2 (Rev. 05-01)

787
CS


INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
04/07/2011 10:23:32

154

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 10:24:00 |
| Cal Check | 0.101 | 10:24:18 |
| Air Blank | 0.000 | 10:24:48 |



Operator's Signature