



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4
5.10.2011

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 05/05/11 Cal 1 0.101 AC Cal 2 0.101 AC

Between the dates of 04/07/2011 and 05/05/2011 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A
The instrument was returned to service on N/A


Cheryl K. Straub (handwritten signature)

Cheryl K. Straub

05/10/2011 (handwritten date)

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 10 day of May, 20 11.



Sue A. Soza (handwritten signature)
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

177 CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY
5.10.2011

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department
DATE 05/05/2011 TIME 10:17
INTOXILYZER SERIAL # 80-002345 LOCATION CRIME LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 101 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Cheryl Straub

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

277 CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
05/05/2011

10:19:28

BH

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:20:18
Diagnostic Test	Pass	10:20:45
Air Blank	0.000	10:21:12
0.100 Cal Check	0.101	10:21:31
Air Blank	0.000	10:22:01
Subject Test	0.000	10:22:33
Air Blank	ABT*	10:22:40
Air Blank	0.000	10:23:08

*Sequence Aborted

Not a Successfully Completed Test Sequence

387
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
05/05/2011 10:24:20

BY

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:24:51
Air Blank	IPS**	10:24:59

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7/7
08

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
05/05/2011

10:25:24

B4

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:26:13
Diagnostic Test	Pass	10:26:40
Air Blank	0.000	10:27:08
0.100 Cal Check	0.101	10:27:27
Air Blank	0.000	10:27:56
Subject Test	0.087*	10:31:04
Air Blank	0.000	10:31:36
Air Blank	0.000	10:32:03
Subject Test	XXX**	10:32:18
Air Blank	0.000	10:32:50

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

577
CS


INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
05/05/2011 10:33:15

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

ASG

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:33:42
Cal Check	0.101	10:34:01
Air Blank	0.000	10:34:30



Operator's Signature

EXHIBIT G-2

687
05

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

BY
5.10.2011

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
DATE 05/05/2011 TIME 10:17
Intoxilyzer Serial # 80-002345 Location CRIME LAB

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

787
CS


154

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
05/05/2011 10:17:31

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:17:58
Cal Check	0.101	10:18:17
Air Blank	0.000	10:18:47



Operator's Signature