



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

By  
9.27.2011

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 9-23-11 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 8-25-11 and 9-23-11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.  
The instrument was returned to service on N/A.


Clare Conley

Clare L. Conley

9.23.11

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 23, day of September, 2011.



Sue A Soza

Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

1067  
en

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

BSY  
9-27-2011

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clara Conley AGENCY Tucson Police Department

DATE 9-23-11 TIME 1113

INTOXILYZER SERIAL # 80-002345 LOCATION ODD

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.

(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.

(✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed

(✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.

(✓) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 099 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS None.  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Clara Conley

2077  
Cm

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
09/23/2011

11:15:07

TS4

---

Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772

---

QAS: C. CONLEY #34517  
TUCSON PD

---

Operator: C. CONLEY #34517  
TUCSON PD

---

Subject: TEST, SQAP  
DOB: 11/11/1111  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	11:15:51
Diagnostic Test	Pass	11:16:18
Air Blank	0.000	11:16:45
0.100 Cal Check	0.100	11:17:04
Air Blank	0.000	11:17:34
Subject Test	0.000	11:17:50
Air Blank	ABT*	11:17:56
Air Blank	0.000	11:18:25

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
09/23/2011 11:18:29

3867  
Cm  
B4

---

Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772

---

QAS: C. CONLEY #34517  
TUCSON PD

---

Operator: C. CONLEY #34517  
TUCSON PD

---

Subject: TEST, SQAP  
DOB: 11/11/1111  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	11:18:54
Air Blank	IPS**	11:19:02

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

4867  
CW

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number : 80-002345  
09/23/2011 11:19:09

BSH

Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772

QAS: C. CONLEY #34517  
TUCSON PD

Operator: C. CONLEY #34517  
TUCSON PD

Subject: TEST, SQAP  
DOB: 11/11/1111  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:19:56
Diagnostic Test	Pass	11:20:23
Air Blank	0.000	11:20:51
0.100 Cal Check	0.100	11:21:09
Air Blank	0.000	11:21:39
Subject Test	0.278*	11:24:47
Air Blank	0.000	11:25:21
Air Blank	0.000	11:25:48
Subject Test	XXX**	11:26:17
Air Blank	0.000	11:26:51

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number: 80-002345  
09/23/2011

11:27:07

5067  
Cm

-----  
Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772  
-----

TS4

-----  
QAS: C. CONLEY #34517  
TUCSON PD  
-----

---

Test	g/210L	Time
Air Blank	0.000	11:27:35
Cal Check	0.099	11:27:54
Air Blank	0.000	11:28:24

*Claire Conley*

\_\_\_\_\_  
Operator's Signature

EXHIBIT G-2

6587  
cu

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

T54  
9-27-2011

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley AGENCY Tucson Police Department  
DATE 9-23-11 TIME 1113  
Intoxilyzer Serial # 80-002345 Location 0DD

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 100 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: ARIZONA  
Serial Number: 80-002345  
09/23/2011 11:13:46

7567  
Cm

-----  
Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772  
-----

TS4

-----  
QAS: C. CONLEY #34517  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	11:14:13
Cal Check	0.100	11:14:32
Air Blank	0.000	11:15:02

*Clare Conley*

\_\_\_\_\_  
Operator's Signature