



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 270 S. Stone Avenue Tucson, AZ 85701-1917

SC
 11/21/11

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/17/11 Cal 1 0.101 AC Cal 2 0.099 AC

Between the dates of 10/21/11 and 11/17/11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
 The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub


11/21/11

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 21 day of November, 2011.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

177
CS

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

SC
11/24/11

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 11/17/11 TIME 08:09

INTOXILYZER SERIAL # 80-002345 LOCATION TEAM 5

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Cheryl Straub

287
C9

SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
11/17/2011 08:16:20

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:17:23
Diagnostic Test	Pass	08:17:50
Air Blank	0.000	08:18:18
0.100 Cal Check	0.102	08:18:36
Air Blank	0.000	08:19:06
Subject Test	0.000	08:19:45
Air Blank	ABT*	08:19:51
Air Blank	0.000	08:20:20

*Sequence Aborted

Not a Successfully Completed Test Sequence

377
CS

SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
11/17/2011

08:20:24

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TPD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:21:14
Air Blank	IPS**	08:21:21

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

477
CS

SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
11/17/2011 08:21:26

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:22:17
Diagnostic Test	Pass	08:22:44
Air Blank	0.000	08:23:12
0.100 Cal Check	0.101	08:23:31
Air Blank	0.000	08:24:00
Subject Test	0.162*	08:27:09
Air Blank	0.000	08:27:41
Air Blank	0.000	08:28:09
Subject Test	XXX**	08:28:38
Air Blank	0.000	08:29:12

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

577
es

SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
11/17/2011 08:29:30

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:29:57
Cal Check	0.099	08:30:16
Air Blank	0.000	08:30:45



Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

697
cg
SC
11/17/11

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
DATE 11/17/11 TIME 08:09
Intoxilyzer Serial # 80-002345 Location TEAM 5

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution W/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

777
CS

50

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
11/17/2011 08:09:54

Standard Lot#: 20910100A1
Last Changed By: C. STRAUB #50772

OAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:10:22
Cal Check	0.101	08:10:40
Air Blank	0.000	08:11:10

C. Straub
Operator's Signature