



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

sc
1/7/13

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12-13-2012 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 11-15-2012 and 12-13-2012 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on _____
 The instrument was returned to service on _____

Terry Gallegos
 Terry S. Gallegos


1-3-2013

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 3rd, day of January, 2013.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
B4

SC
1/7/13

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 12-13-2012 TIME 1213

INTOXILYZER SERIAL # 80-002345 LOCATION ODD

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Terry Gallegos

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
12/13/2012

12:15:26

20f7
B4

Standard Lot#: 23511100A2
Last Changed By: T. GALLEGOS #28847

32

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:16:19
Diagnostic Test	Pass	12:16:46
Air Blank	0.000	12:17:14
0.100 Cal Check	0.100	12:17:32
Air Blank	0.000	12:18:02
Subject Test	0.000	12:18:40
Air Blank	0.000	12:19:07
Five Minute Wait	ABT*	
Air Blank	0.000	12:19:43

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
12/13/2012

12:19:54

30f7
TSh

5C

Standard Lot#: 23511100A2
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:20:52
Diagnostic Test	Pass	12:21:20
Air Blank	RFI*	12:21:30
Air Blank	IPS**	12:21:38

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002345
12/13/2012

12:21:53

40F7
TS4

Standard Lot#: 23511100A2
Last Changed By: T. GALLEGOS #28847

5c

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:22:47
Diagnostic Test	Pass	12:23:14
Air Blank	0.000	12:23:41
0.100 Cal Check	0.100	12:24:00
Air Blank	0.000	12:24:30
Subject Test	0.448*	12:27:38
Air Blank	0.000	12:28:11
Air Blank	0.000	12:28:39
Subject Test	XXX**	12:29:08
Air Blank	0.000	12:29:41

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
12/13/2012 12:30:01


Sof 7
Bh

Standard Lot#: 23511100A2
Last Changed By: T. GALLEGOS #28847

5c

QAS: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:30:28
Cal Check	0.099	12:30:47
Air Blank	0.000	12:31:16



Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6087
TS4

5C
1/7/13

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 12-13-2012 TIME 1213
Intoxilyzer Serial # 80-002345 Location DDD

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002345
12/13/2012 12:13:56

70f7
TS4

Standard Lot#: 23511100A2
Last Changed By: T. GALLEGOS #28847

5C

QAS: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:14:23
Cal Check	0.100	12:14:42
Air Blank	0.000	12:15:11



Operator's Signature