



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

SC  
2/14/14

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 2.6.2014 Cal 1 0.099 AC Cal 2 0.099 AC

Between the dates of 1.9.2014 and 2.6.2014 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.  
The instrument was returned to service on —.

Terry Gallegos

Terry S. Gallegos


2.13.2014

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 13<sup>th</sup>, day of February, 2014.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7  
TGH  
SC  
2/14/14

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 2.6.2014 TIME 1217

INTOXILYZER SERIAL # 80-002345 LOCATION ODD

(  ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(  ) 1. Clock time check.

(  ) 2. Date check.

OPERATIONAL TESTS

(  ) 1. Alcohol-free subject test result 0. 000 AC.

(  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed

(  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.

(  ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 099 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE

Terry Gallegos

INTOXILYZER 8000

Location:

Serial Number: 80-002345

Core Version: 8105.48

02/06/2014

12:19:20

20 of 7  
T04  
SC

---

Standard Lot#: 25913100A2

Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847

TUCSON PD

---

Operator: T. GALLEGOS #28847

TUCSON PD

---

Subject: TEST,SQAP

DOB: 11/11/1987

Sex: F

Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:20:13
Diagnostic Test	Pass	12:20:48
Air Blank	0.000	12:21:15
0.100 Cal Check	0.100	12:21:36
Air Blank	0.000	12:22:05
Subject Test	0.000	12:22:22
Air Blank	0.000	12:22:49
Wait	ABT*	
Air Blank	0.000	12:23:24

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location:  
Serial Number: 80-002345  
Core Version: 8105.48  
02/06/2014

12:23:27

3 of 7  
Dy  
sc

---

Standard Lot#: 25913100A2  
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847  
TUCSON PD

---

Operator: T. GALLEGOS #28847  
TUCSON PD

---

Subject: TEST,SQAP  
DOB: 11/11/1987  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:24:19
Diagnostic Test	Pass	12:24:54
Air Blank	0.000	12:25:21
0.100 Cal Check	RFI*	12:25:36
Air Blank	IPS**	12:25:44

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location:  
Serial Number: 80-002345  
Core Version: 8105.48  
02/06/2014

12:25:50

4 of 7  
156

5C

---

Standard Lot#: 25913100A2  
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847  
TUCSON PD

---

Operator: T. GALLEGOS #28847  
TUCSON PD

---

Subject: TEST, SQAP  
DOB: 11/11/1987  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:26:39
Diagnostic Test	Pass	12:27:14
Air Blank	0.000	12:27:41
0.100 Cal Check	0.099	12:28:02
Air Blank	0.000	12:28:32
Subject Test	0.070*	12:31:39
Air Blank	0.000	12:32:11
Air Blank	0.000	12:32:38
Subject Test	XXX**	12:33:10
Air Blank	0.000	12:33:44

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location:

Serial Number: 80-002345

Core Version: 8105.48

02/06/2014

12:34:13

5077  
TSh

-----  
Standard Lot#: 25913100A2

Last Changed By: T. GALLEGOS #28847

5C

-----  
QAS: T. GALLEGOS #28847

TUCSON PD

-----  
Operator: T. GALLEGOS #28847

TUCSON PD

-----  
Test                                    g/210L                                    Time  
-----  
Air Blank                                0.000                                    12:34:43  
0.100 Cal Check                        0.099                                    12:35:03  
Air Blank                                0.000                                    12:35:33

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

60f 7  
TS4  
SC  
2/14/14

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department  
DATE 2-6-2014 TIME 1217  
Intoxilyzer Serial # 80-002345 Location ODD

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 099 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000

Location:

Serial Number: 80-002345

Core Version: 8105.48

02/06/2014

12:17:09

*Fof7*  
*TS4*

-----  
Standard Lot#: 25913100A2

Last Changed By: T. GALLEGOS #28847

*SC*

-----  
QAS: T. GALLEGOS #28847

TUCSON PD

-----  
Operator: T. GALLEGOS #28847

TUCSON PD CRIME LAB

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      12:17:37  
0.100 Cal Check              0.099                      12:17:57  
Air Blank                      0.000                      12:18:27