



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
5-10-2011

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 05/05/11 Cal 1 0.098 AC Cal 2 0.097 AC

Between the dates of 04/07/2011 and 05/05/2011 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A
The instrument was returned to service on N/A


Cheryl K. Straub (handwritten signature)

Cheryl K. Straub

05/10/2011 (handwritten date)

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 10 day of May, 20 11.



Sue A. Soza (handwritten signature)

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

187 CG

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TS4
5.10-2011

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 05/05/2011 TIME 05:51

INTOXILYZER SERIAL # 80-002759 LOCATION TEAM 4

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Clyde Str

287
CS

INTOXILYZER 8000

Location:

Serial Number : 80-002759

05/05/2011

05:53:11

TSK

Standard Lot#: 719919

Last Changed By: C. CONLEY

#34517

QAS: C. STRAUB

#50772

TUCSON PD

Operator: C. STRAUB

#50772

TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	05:54:03
Diagnostic Test	Pass	05:54:31
Air Blank	0.000	05:54:58
0.100 Cal Check	0.098	05:55:17
Air Blank	0.000	05:55:47
Subject Test	0.000	05:56:02
Air Blank	ABT*	05:56:09
Air Blank	0.000	05:56:37

*Sequence Aborted

Not a Successfully Completed Test Sequence

377
CS

15h

INTOXILYZER 8000
 Location:
 Serial Number : 80-002759
 05/05/2011 05:56:41

Standard Lot#: 719919
 Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
 TUCSON PD

Operator: C. STRAUB #50772
 TUCSON PD

Subject: SQAP, SQAP
 DOB: 11/11/1111
 Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	05:57:11
Air Blank	IPS**	05:57:19

*RFI Detect
 **Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

487
CS

INTOXILYZER 8000

Location:

Serial Number : 80-002759

05/05/2011

05:57:25

BSH

Standard Lot#: 719919

Last Changed By: C. CONLEY

#34517

QAS: C. STRAUB

#50772

TUCSON PD

Operator: C. STRAUB

#50772

TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	05:58:20
Diagnostic Test	Pass	05:58:47
Air Blank	0.000	05:59:15
0.100 Cal Check	0.099	05:59:33
Air Blank	0.000	06:00:03
Subject Test	0.171*	06:03:11
Air Blank	0.000	06:03:44
Air Blank	0.000	06:04:11
Subject Test	XXX**	06:04:44
Air Blank	0.000	06:05:18

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

587
02

INTOXILYZER 8000

Location:

Serial Number: 80-002759

05/05/2011

06:05:41

BSH

Standard Lot#: 719919

Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772

TUCSON PD

Test	g/210L	Time
Air Blank	0.000	06:06:08
Cal Check	0.097	06:06:27
Air Blank	0.000	06:06:57



Operator's Signature

697
CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

TS4
5.10.2011

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
DATE 05/05/2011 TIME 05:51
Intoxilyzer Serial # 80-002759 Location TEAM 4

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

787
08


INTOXILYZER 8000
Location:
Serial Number: 80-002759
05/05/2011 05:51:48

PH

Standard Lot#: 719919
Last Changed By: C. CONLEY #34517

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	05:52:16
Cal Check	0.098	05:52:35
Air Blank	0.000	05:53:04



Operator's Signature