



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

SC  
11/21/11

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/17/11 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 10/21/11 and 11/17/11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A

The instrument was returned to service on N/A

Cheryl K. Straub

Cheryl K. Straub


11/21/11

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 21 day of November, 2011.

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

11/17/11

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

SC  
11/21/11

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl Straub AGENCY Tucson Police Department  
DATE 11/17/11 TIME 05:37  
INTOXILYZER SERIAL # 80-002759 LOCATION TEAM 4

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 097 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS INVALID SAMPLE TEST WAS REPEATED. INITIAL  
TEST RECORDED AS SUBJECT TEST.

SIGNATURE Cheryl Straub

287  
CS

SC

INTOXILYZER 8000

Location:

Serial Number : 80-002759

11/17/2011

05:39:00

---

Standard Lot#: 04911100A1

Last Changed By: C. STRAUB

#50772

---

QAS: C. STRAUB

#50772

TUCSON PD

---

Operator: C. STRAUB

#50772

TUCSON PD

---

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	05:39:54
Diagnostic Test	Pass	05:40:22
Air Blank	0.000	05:40:49
0.100 Cal Check	0.097	05:41:08
Air Blank	0.000	05:41:37
Subject Test	0.000	05:41:59
Air Blank	ABT*	05:42:06
Air Blank	0.000	05:42:35

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

387  
C8

5C

INTOXILYZER 8000

Location:

Serial Number : 80-002759

11/17/2011

05:42:43

Standard Lot#: 04911100A1

Last Changed By: C. STRAUB

#50772

QAS: C. STRAUB

#50772

TUCSON PD

Operator: C. STRAUB

#50772

TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	05:43:25
Air Blank	IPS**	05:43:33

\*RFI Detect

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

427  
CS

SK

INTOXILYZER 8000  
Location:  
Serial Number : 80-002759  
11/17/2011 05:43:42

Standard Lot#: 04911100A1  
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772  
TUCSON PD

Operator: C. STRAUB #50772  
TUCSON PD

Subject: SQAP, SQAP  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	05:44:31
Diagnostic Test	Pass	05:44:58
Air Blank	0.000	05:45:25
0.100 Cal Check	0.098	05:45:44
Air Blank	0.000	05:46:14
Subject Test	0.129*	05:49:22
Air Blank	0.000	05:49:55
Air Blank	0.000	05:50:22
Subject Test	0.197	05:51:16
Air Blank	0.000	05:51:49
Five Minute Wait		
Air Blank	0.000	05:56:44
Subject Test	XXX**	05:57:14
Air Blank	0.000	05:57:48

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

577  
C8


SC

INTOXILYZER 8000  
Location:  
Serial Number: 80-002759  
11/17/2011 05:58:26

-----  
Standard Lot#: 04911100A1  
Last Changed By: C. STRAUB #50772  
-----

QAS: C. STRAUB #50772  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	05:58:54
Cal Check	0.097	05:59:12
Air Blank	0.000	05:59:42



\_\_\_\_\_  
Operator's Signature

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

697  
C8

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

SC  
11/21/11

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department  
DATE 11/17/11 TIME 05:37  
Intoxilyzer Serial # 80-002759 Location TEAM 4

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 097 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

787  
CS

50

INTOXILYZER 8000

Location:

Serial Number: 80-002759

11/17/2011

05:37:32


-----  
Standard Lot#: 04911100A1

Last Changed By: C. STRAUB #50772

-----  
QAS: C. STRAUB #50772

TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	05:37:59
Cal Check	0.097	05:38:18
Air Blank	0.000	05:38:47

  
-----  
Operator's Signature