



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

SC
4/11/12

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4-5-2012 Cal 1 0.097 AC Cal 2 0.103 AC

Between the dates of 3-8-2012 and 4-5-2012 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Terry S. Gallegos

Terry S. Gallegos


4.11.2012

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 11th, day of April, 2012.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 9
TS4
SC
4/11/12

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 4-5-2012 TIME 1906
INTOXILYZER SERIAL # 80-002759 LOCATION ODE

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning. ✓✓
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 103 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Terry Gallegos

INTOXILYZER 8000

Location:

Serial Number : 80-002759

04/05/2012

19:08:10

20 of 9
TS4

Standard Lot#: 04911100A1

Last Changed By: C. STRAUB

#50772

SC

QAS: T. GALLEGOS

#28847

TUCSON PD

Operator: T. GALLEGOS

#28847

TUCSON PD

Subject: TEST, SQAP

DOB: 11/11/1987

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:09:04
Diagnostic Test	Pass	19:09:31
Air Blank	0.000	19:09:58
0.100 Cal Check	0.099	19:10:17
Air Blank	0.000	19:10:47
Subject Test	0.000	19:11:04
Air Blank	ABT*	19:11:13
Air Blank	0.000	19:11:41

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location:

Serial Number : 80-002759

04/05/2012

19:11:47

30f9
TS4

Standard Lot#: 04911100A1

Last Changed By: C. STRAUB

#50772

SC

QAS: T. GALLEGOS

#28847

TUCSON PD

Operator: T. GALLEGOS

#28847

TUCSON PD

Subject: TEST, SQAP

DOB: 11/11/1987

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:12:40
Diagnostic Test	Pass	19:13:07
Air Blank	RFI*	19:13:20
Air Blank	IPS**	19:13:29

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location:

Serial Number : 80-002759

04/05/2012

19:13:38

40f 9
734

Standard Lot#: 04911100A1

Last Changed By: C. STRAUB

#50772

QAS: T. GALLEGOS

#28847

TUCSON PD

Operator: T. GALLEGOS

#28847

TUCSON PD

30

Subject: TEST, SQAP

DOB: 11/11/1987

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:14:27
Diagnostic Test	Pass	19:14:55
Air Blank	0.000	19:15:22
0.100 Cal Check	0.097	19:15:41
Air Blank	0.000	19:16:10
Subject Test	0.223*	19:19:19
Air Blank	0.000	19:19:51
Air Blank	0.000	19:20:18
Subject Test	XXX**	19:20:52
Air Blank	0.000	19:21:25

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location:

Serial Number: 80-002759

04/05/2012

19:21:44

50f9
T54

Standard Lot#: 04911100A1


Last Changed By: C. STRAUB #50772

QAS: T. GALLEGOS #28847

TUCSON PD

5C

Test	g/210L	Time
Air Blank	0.000	19:22:11
Cal Check	0.103	19:22:30
Air Blank	0.000	19:22:59



Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

60f9
TS4 SC
4/11/12

QA Specialist Terry Gallegas 28847 AGENCY Tucson Police Department
DATE 4.5.2012 TIME 1906
Intoxilyzer Serial # 80-002759 Location ODE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegas

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000

Location:

Serial Number: 80-002759

04/05/2012

19:06:00

7059
TS4

Standard Lot#: 04911100A1

Last Changed By: C. STRAUB #50772


SC

QAS: T. GALLEGOS #28847

TUCSON PD

Test g/210L Time

Air Blank 0.000 19:06:27
Cal Check 0.097 19:06:46
Air Blank 0.000 19:07:16



Operator's Signature



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

8079
TS4
SC
4/11/12

Date: 4.5.2012
 QAS: Terry Gallegos 28847
 Intoxilyzer Serial Number: 80-002759
 New Lot Number: 23511100A2
 New Expiration Date: 11.1.2013

- (✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 2. Remove the old gas tank and install the new one.
- (✓) 3. Hit the "ESC" key twice and swipe the QAS card.
- (✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change."
 Select "D" for dry gas and enter the following information:
 Enter Target Value: 0.100 ✓
 Enter Control Lot #: From the tank label
 Enter Exp Date: From the tank label (mm/dd/yyyy)
 Edit info Y/N? Y or N
 Update Std Info? Y ENTER
 "Please wait saving..."
- (✓) 5. Hit the "ESC" key twice.
- (✓) 6. Go to the "M-Maintenance" menu and select "Control Testing."
 Select "D" for dry control test. After selection is made press ENTER.
- (✓) 7. Air blank completed.
- (✓) 8. Calibration check completed. Test results 0. 100 AC.
- (✓) 9. Air blank completed.
- (✓) 10. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

