



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

sc  
1/7/13

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12-13-2012 Cal 1 0.098 AC Cal 2 0.097 AC

Between the dates of 11-15-2012 and 12-13-2012 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on                     .  
The instrument was returned to service on                     .

Terry Gallegos

Terry S. Gallegos

1.3.2013


ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 3rd, day of January, 2013.

Edma Escarcega

Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5



Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

10f7  
TS4

SC  
1/7/13

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 12-13-2012 TIME 0915

INTOXILYZER SERIAL # 80-002759 LOCATION ODE

(  ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.
- (  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.
- (  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (  ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 097 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

---



---



---



---

SIGNATURE Terry Gallegos

INTOXILYZER 8000  
Location:  
Serial Number : 80-002759  
12/13/2012

09:19:41

2 of 7  
TS4

SL

Standard Lot#: 23511100A2  
Last Changed By: T. GALLEGOS #28847

OAS: T. GALLEGOS #28847  
TUCSON PD

Operator: T. GALLEGOS #28847  
TUCSON PD

Subject: TEST, SQAP

DOB: 11/11/1987

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:20:34
Diagnostic Test	Pass	09:21:02
Air Blank	0.000	09:21:29
0.100 Cal Check	0.098	09:21:48
Air Blank	0.000	09:22:17
Subject Test	0.000	09:22:48
Air Blank	0.000	09:23:15
Five Minute Wait	ABT*	
Air Blank	0.000	09:23:49

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location:

Serial Number : 80-002759

12/13/2012

09:25:03

3057  
T34

Standard Lot#: 23511100A2

Last Changed By: T. GALLEGOS

#28847

SL

OAS: T. GALLEGOS

#28847

TUCSON PD

Operator: T. GALLEGOS

#28847

TUCSON PD

Subject: TEST, SOAP

DOB: 11/11/1987

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:25:53
Diagnostic Test	Pass	09:26:20
Air Blank	RFI*	09:26:33
Air Blank	IPS**	09:26:41

\*RFI Detect

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location:

Serial Number : 80-002759

12/13/2012

09:26:48

4057  
TJ4

Standard Lot#: 23511100A2

Last Changed By: T. GALLEGOS

#28847

SL

QAS: T. GALLEGOS

#28847

TUCSON PD

Operator: T. GALLEGOS

#28847

TUCSON PD

Subject: TEST, SQAP

DOB: 11/11/1987

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:27:37
Diagnostic Test	Pass	09:28:05
Air Blank	0.000	09:28:32
0.100 Cal Check	0.097	09:28:51
Air Blank	0.000	09:29:20
Subject Test	0.226*	09:32:29
Air Blank	0.000	09:33:01
Air Blank	0.000	09:33:29
Subject Test	XXX**	09:33:58
Air Blank	0.000	09:34:32

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location:

Serial Number: 80-002759

12/13/2012

09:35:32

5077  
TS4

-----  
Standard Lot#: 23511100A2

Last Changed By: T. GALLEGOS #28847

-----  
QAS: T. GALLEGOS #28847

TUCSON PD  
-----

SC

Test	g/210L	Time
Air Blank	0.000	09:36:00
Cal Check	0.097	09:36:18
Air Blank	0.000	09:36:48

  
-----  
Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6077  
TS4  
1/7/13

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department  
DATE 12.13.2012 TIME 0915  
Intoxilyzer Serial # 80-002759 Location ODE

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 098 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000

Location:

Serial Number: 80-002759

12/13/2012

09:15:56

7057  
754

-----  
Standard Lot#: 23511100A2

Last Changed By: T. GALLEGOS #28847

-----  
QAS: T. GALLEGOS #28847

TUCSON PD

SC

-----

Test	g/210L	Time
Air Blank	0.000	09:16:23
Cal Check	0.098	09:16:42
Air Blank	0.000	09:17:11

-----



Operator's Signature