



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4
11-25-2015

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/12/2015 Cal 1 0.097 AC Cal 2 0.095 AC

Between the dates of 10/14/2015 and 11/12/2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on _____.
The instrument was returned to service on _____.

William L Loomis

William L. Loomis


11/25/2015

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 25th, day of November, 2015.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
WLL
TSG 11-25-2015

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department

DATE 11/12/2015 TIME 15:07

INTOXILYZER SERIAL # 80-002759 LOCATION ODM -SGT'S OFFICE

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 095 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Will Loomis

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
11/12/2015 15:10:20

2 of 7
WLL
T84

Standard Lot#: 11514100A2
Last Changed By: S. CIONE #50616

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:11:17
Diagnostic Test	Pass	15:11:54
Air Blank	0.000	15:12:21
0.100 Cal Check	0.096	15:12:42
Air Blank	0.000	15:13:12
Subject Test	0.000	15:13:29
Air Blank	0.000	15:13:57
Wait	ABT*	
Air Blank	0.000	15:14:32

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
11/12/2015 15:15:25

3 of 7
6 WLL
TS4

Standard Lot#: 1151410192
Last Changed By: S. CIONE #50616

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:16:26
Diagnostic Test	Pass	15:17:03
Air Blank	0.000	15:17:30
0.100 Cal Check	RFI*	15:17:46
Air Blank	IPS**	15:17:55

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
11/12/2015 15:10:30

4 of 7
WLL
TS4

Standard Lot#: 11514100A2
Last Changed By: S. CIONE #50616

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:19:38
Diagnostic Test	Pass	15:20:15
Air Blank	0.000	15:20:42
0.100 Cal Check	0.097	15:21:03
Air Blank	0.000	15:21:33
Subject Test	0.273*	15:24:40
Air Blank	0.000	15:25:14
Air Blank	0.000	15:25:41
Subject Test	XXX**	15:25:54
Air Blank	0.000	15:26:29

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
0 WLL
T84

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
11/12/2015 15:27:32

Standard Lot#: 11514100A2
Last Changed By: S. CIONE #50616

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	15:28:03
0.100 CaI Check	0.095	15:28:24
Air Blank	0.000	15:28:54

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
WLL
BL
11-25-2015

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 11/12/2015 TIME 15:07
Intoxilyzer Serial # 80-002759 Location ODM - SGT'S OFFICE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE William Loomis

DPS Form Exh G-2 (Rev. 05-01)

7 of 7
0 WLL

TS4

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
11/12/2015 15:07:39

Standard Lot#: 11514100A2
Last Changed By: S. CLONE #50616

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	15:08:08
0.100 Cal Check	0.097	15:08:29
Air Blank	0.000	15:08:59

Preventative Maintenance
Performed -- Timer Reset