



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

T84  
1.13.2016

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1/7/2016 Cal 1 0.096 AC Cal 2 0.094 AC

Between the dates of 12/10/2015 and 1/7/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on \_\_\_\_\_.  
The instrument was returned to service on \_\_\_\_\_.

William L Loomis

William L. Loomis


1/12/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 12<sup>th</sup>, day of January, 2016.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7  
WLL  
T84, 1.13.2016

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department

DATE 1/7/2016 TIME 10:27

INTOXILYZER SERIAL # 80-002759 LOCATION ODM-SGT'S OFFICE

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 094 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

---



---



---



---

SIGNATURE William Loomis

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
01/07/2016 10:29:24

2 of 7  
WLL  
TS4

Standard Lot#: 11514100A2  
Last Changed By: S. CLONE #50616

OAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SOAP, SOAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:30:19
Diagnostic Test	Pass	10:30:56
Air Blank	0.000	10:31:23
0.100 Cal Check	0.096	10:31:44
Air Blank	0.000	10:32:13
Subject Test	0.000	10:32:31
Air Blank	0.000	10:32:58
Wait	ABT*	
Air Blank	0.000	10:33:33

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

3 of 7  
will  
TS4

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
01/07/2016 10:34:41

Standard Lot#: 11514100A2  
Last Changed By: S. CIONE #50616

OAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SOAP, SOAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:35:39
Diagnostic Test	Pass	10:36:16
Air Blank	0.000	10:36:44
0.100 Cal Check	RFI*	10:36:57
Air Blank	IPS**	10:37:05

\*RFI Detect  
\*\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
01/07/2016 10:37:57

4 of 7  
wru  
TS4

Standard Lot#: 11514100A2  
Last Changed By: S. CLONE #50616

OAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SOAP, SOAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:38:50
Diagnostic Test	Pass	10:39:28
Air Blank	0.000	10:39:55
0.100 Cal Check	0.096	10:40:16
Air Blank	0.000	10:40:45
Subject Test	0.229*	10:43:53
Air Blank	0.000	10:44:26
Air Blank	0.000	10:44:54
Subject Test	XXX**	10:45:11
Air Blank	0.000	10:45:46

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

5 of 7  
WU  
TSL

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
01/07/2016 10:46:44

-----  
Standard Lot#: 11514100A2  
Last Changed By: S. CLONE #50616  
-----

QAS: W. LOOMIS #53545  
TUCSON PD  
-----

Operator: W. LOOMIS #53545  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	10:47:15
0.100 Cal Check	0.094	10:47:36
Air Blank	0.000	10:48:05

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7  
will  
TS4 1.13.2016

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department  
DATE 1/7/2016 TIME 10:27  
Intoxilyzer Serial # 80-002759 Location ODM - SGT'S OFFICE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Willie Loomis

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
01/07/2016 10:27:04

7 of 7  
WLL  
TSK

Standard Lot#: 11514100A2  
Last Changed By: S. CIONE #50616

OAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:27:33
0.100 Cal Check	0.096	10:27:54
Air Blank	0.000	10:28:24

Preventative Maintenance  
Performed -- Timer Reset