



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4
6-6-2016

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5/26/2016 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 5/3/2016 and 5/26/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on
The instrument was returned to service on

William L Loomis

William L. Loomis


6/1/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 1 June, 2016, day of

A.M. Hernandez

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7  
WLL  
TS4  
6.6.2016

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department

DATE 5/26/2016 TIME 09:04

INTOXILYZER SERIAL # 80-002759 LOCATION ODM-SGT'S OFFICE

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 098 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Will Loomis

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7  
WLL  
BY

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
05/26/2016 09:07:23

Standard Lot#: 06616100A1  
Last Changed By: W. LOOMIS #53545

OAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SOAP, SOAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:08:15
Diagnostic Test	Pass	09:08:53
Air Blank	0.000	09:09:20
0.100 Cal Check	0.099	09:09:41
Air Blank	0.000	09:10:11
Subject Test	0.000	09:10:29
Air Blank	0.000	09:10:56
Wait	ABT*	
Air Blank	0.000	09:11:30

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

3 of 7  
WLL  
TS4

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
05/26/2016 09:12:03

Standard Lot#: 06616100A1  
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SOAP, SOAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:12:55
Diagnostic Test	Pass	09:13:33
Air Blank	0.000	09:14:00
0.100 Cal Check	RFI*	09:14:13
Air Blank	IPS**	09:14:21

\*RFI Detect  
\*\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

4 of 7  
WLL  
B4

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
05/26/2016 09:14:42

Standard Lot#: 06616100A1  
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SOAP, SOAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:15:43
Diagnostic Test	Pass	09:16:20
Air Blank	0.000	09:16:47
0.100 Cal Check	0.099	09:17:08
Air Blank	0.000	09:17:38
Subject Test	0.345*	09:20:45
Air Blank	0.000	09:21:19
Air Blank	0.000	09:21:46
Subject Test	XXX**	09:22:07
Air Blank	0.000	09:22:42

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
05/26/2016 09:23:18

5 of 7  
TS4  
WV

Standard Lot#: 06616100A1  
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:23:47
0.100 Cal Check	0.098	09:24:08
Air Blank	0.000	09:24:38

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7  
WLL  
TSY  
6-6-2016

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department  
DATE 5/26/2016 TIME 09:04  
Intoxilyzer Serial # 90-002759 Location ODM - SGT'S OFFICE

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 099 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Willie Loomis

7 of 7  
D will  
B4

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-002759  
Core Version: 8105.48  
05/26/2016 09:04:43

Standard Lot#: 06616100A1  
Last Changed By: W. LOOMIS #53545

GAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:05:13
0.100 Cal Check	0.099	09:05:34
Air Blank	0.000	09:06:03

Preventative Maintenance  
Performed -- Timer Reset