



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002759 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSy
7-26-2016

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7/21/2016 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 6/23/2016 and 7/21/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on
The instrument was returned to service on

Handwritten signature of William L. Loomis

William L. Loomis


7/21/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 21st, day of July, 2016.

Handwritten signature of Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
Will
By 7-26-2016

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 7/21/2016 TIME 09:08
INTOXILYZER SERIAL # 80-002759 LOCATION ODM-SGT'S OFFICE

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Will Loomis

2 of 7
WLL
TS4

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
07/21/2016 09:10:24

Standard Lot#: 06616100A1
Last Changed By: W. LOOMIS #53545

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, SOAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:11:27
Diagnostic Test	Pass	09:12:04
Air Blank	0.000	09:12:31
0.100 Cal Check	0.097	09:12:52
Air Blank	0.000	09:13:22
Subject Test	0.000	09:13:37
Air Blank	0.000	09:14:04
Wait	ABT*	
Air Blank	0.000	09:14:40

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7
6 WLL
TS4

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
07/21/2016 09:15:07

Standard Lot#: 06616100A1
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, SOAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:15:59
Diagnostic Test	Pass	09:16:36
Air Blank	0.000	09:17:03
0.100 Cal Check	RFI*	09:17:19
Air Blank	IPS**	09:17:26

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
07/21/2016 09:17:49

4 of 7
WLL
TS4

Standard Lot#: 06616100A1
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, SOAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:18:55
Diagnostic Test	Pass	09:19:32
Air Blank	0.000	09:19:59
0.100 Cal Check	0.097	09:20:20
Air Blank	0.000	09:20:50
Subject Test	0.189*	09:23:57
Air Blank	0.000	09:24:30
Air Blank	0.000	09:24:57
Subject Test	XXX**	09:25:16
Air Blank	0.000	09:25:51

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
WLL
BS

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
07/21/2016 09:26:37

Standard Lot#: 06616100A1
Last Changed By: W. LOOMIS #53545

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:27:08
0.100 Cal Check	0.097	09:27:29
Air Blank	0.000	09:27:58

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
WLL
TS4
7-26-2016

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 7/21/2016 TIME 09:09
Intoxilyzer Serial # 80-002759 Location ODM-SGT'S OFFICE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Will Loomis

DPS Form Exh G-2 (Rev. 05-01)

7 of 7
WLL
TB4

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002759
Core Version: 8105.48
07/21/2016 09:08:45

Standard Lot#: 06616100A1
Last Changed By: W. LOOMIS #53545

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:09:16
0.100 Cal Check	0.097	09:09:37
Air Blank	0.000	09:10:06

Preventative Maintenance
Performed -- Timer Reset