CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002760 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program. TM 1.19.2010

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1-7-10 Cal 1 0.101 AC Cal 2 0.101 AC

Between the dates of 12-10-09 and 1-7-10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

1-14-10

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 14, day of January, 20 10.

OFFICIAL SEAL
MONICA L. BRAVO
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. June 27, 2011

Monica L. Bravo

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

1 of 9

CU

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

JBy 1.19.2010

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 1-7-10 TIME 10:14

INTOXILYZER SERIAL # 80-002760 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.101 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS Regulator failed during SQAP tests (would permit calibration checks but not subject tests). Instrument was taken to the Crime Lab, where the regulator was replaced and remainder of tests performed. Instrument was never taken out of service.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 9
eu

[illegible]
Location: ARIZONA
Serial Number: [illegible]
[illegible] [illegible]

Standard Lot#: [illegible]
Last Changed By: C. CONLEY #34517

CAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SDAP
DOB: [illegible]
Sex: M Weight: [illegible]

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	[illegible]	[illegible]
Diagnostic Test	Pass	[illegible]
Air Blank	[illegible]	[illegible]
0.100 Cal Check	[illegible]	[illegible]
Air Blank	[illegible]	[illegible]
Subject Test	[illegible]	[illegible]
Air Blank	ABT*	[illegible]
Air Blank	[illegible]	[illegible]

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 9
CH

[illegible]

[illegible]
Location: [illegible]
Serial Number: [illegible]
[illegible] [illegible]

Standard Lot#: 8260011
Last Changed By: C. CONLEY #34517

DAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SGAP
DOB: [illegible]
Sex: M Weight: [illegible]

15 Minute Deprivation Period? Yes

Test	[illegible]	Time
Air Blank	RPI*	[illegible]
Air Blank	IPS**	[illegible]

*RPI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 9
CU

[illegible]
Location: [illegible]
Serial Number: [illegible]
[illegible]

Standard Lot#: [illegible]
Last Changed By: C. CONLEY #[illegible]

OAS: C. CONLEY #[illegible]
TUCSON PD

Operator: C. CONLEY #[illegible]
TUCSON PD

Subject: TEST,SCAP
DOB: [illegible]
Sex: M Weight: [illegible]

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	[illegible]	[illegible]
Diagnostic Test	Pass	[illegible]
Air Blank	[illegible]	[illegible]
[illegible] Cal Check	[illegible]	[illegible]
Air Blank	[illegible]	[illegible]
Subject Test	[illegible]*	[illegible]
Air Blank	[illegible]	[illegible]
Air Blank	[illegible]	[illegible]
Subject Test	XXX**	[illegible]
Air Blank	[illegible]	[illegible]

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

589
Cu

54

[illegible]
Location: ARIZONA
Serial Number: [illegible]
[illegible] [illegible]

Standard Lot#: [illegible]
Last Changed By: C. CONLEY #34517

DAS: C. CONLEY #34517
TUCSON PD

Test	g/21L	Time
Air Blank	[illegible]	[illegible]
Cal Check	[illegible]	[illegible]
Air Blank	[illegible]	[illegible]

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

6 of 9
CU

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

BY 1.19.2010

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 1-7-10 TIME 10:14

Intoxilyzer Serial # 80-002760 Location Crime Lab

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 9
CM

54

INTOXILIZER 8000
Location: ARIZONA
Serial Number: 80-[illegible]
01/07/2010 10:14:57

Standard Lot#: [illegible]
Last Changed By: C. CONLEY #34517

Opr: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:15:15
Cal Check	0.101	10:15:34
Air Blank	0.000	10:15:53

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

849
CU

DY
1.19.2010

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 1-7-10 TIME 0843

Intoxilyzer Serial # 80-002760 Location ODW DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002760
01/07/2010 08:43:59

989
eu

Standard Lot#: 826001I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:44:26
Cal Check	0.101	08:44:45
Air Blank	0.000	08:45:14

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.