



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
6-7-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002760 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5/26/10 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 4/29/10 and 5/26/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


6/3/10

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 3rd, day of June, 2010.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TS4
6.7.2010

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 5/26/10 TIME 952

INTOXILYZER SERIAL # 80-002760 LOCATION Team 2-DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

2 of 7 sc

py

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002760
05/26/2010

09:54:29

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:55:13
Diagnostic Test	Pass	09:55:40
Air Blank	0.000	09:56:08
0.100 Cal Check	0.101	09:56:26
Air Blank	0.000	09:56:56
Subject Test	0.000	09:57:16
Air Blank	0.000	09:57:43
Five Minute Wait	ABT*	
Air Blank	0.000	09:58:18

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 7 SC

B4

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002760
05/26/2010 09:58:34

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:59:10
Air Blank	IPS**	09:59:19

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

4 of 7 sc

154

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002760
05/26/2010 09:59:25

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:00:17
Diagnostic Test	Pass	10:00:45
Air Blank	0.000	10:01:12
0.100 Cal Check	0.101	10:01:31
Air Blank	0.000	10:02:01
Subject Test	0.103*	10:05:09
Air Blank	0.000	10:05:41
Air Blank	0.000	10:06:08
Subject Test	XXX**	10:06:41
Air Blank	0.000	10:07:14

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

5 of 7 sc

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002760
05/26/2010 10:07:41

154

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:08:08
Cal Check	0.100	10:08:27
Air Blank	0.000	10:08:57

Sarah Cione
Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

By
6-7-2010

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 5/26/10 TIME 952
Intoxilyzer Serial # 80-002760 Location Team 2-DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

7 of 7 sc

154

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002760
05/26/2010 09:52:41

Standard Lot#: 826001I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:53:08
Cal Check	0.101	09:53:27
Air Blank	0.000	09:53:56

Sarah Cione
Operator's Signature