



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

OK
11-30-11

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002760 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11/14/11 Cal 1 0.097 AC Cal 2 0.098 AC

Between the dates of 10/14/11 and 11/14/11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Sarah Cione

Sarah L. Cione


11/16/11

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 16, day of November, 2011.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1 of 7sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

CM
11-30-11

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cone AGENCY Tucson Police Department

DATE 11/14/11 TIME 1110

INTOXILYZER SERIAL # 80-002760 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS Calibration check performed 1 day past 31 days.

SIGNATURE Sarah Cone

2 of 7 sc
cm

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002760
11/14/2011 11:12:02

Standard Lot#: 04911100A1
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:12:45
Diagnostic Test	Pass	11:13:12
Air Blank	0.000	11:13:40
0.100 Cal Check	0.098	11:13:58
Air Blank	0.000	11:14:28
Subject Test	0.000	11:14:56
Air Blank	0.000	11:15:23
Five Minute Wait	ABT*	
Air Blank	0.000	11:16:02

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 7 sc
ln

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002760
11/14/2011 11:16:11

Standard Lot#: 04911100A1
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:16:56
Diagnostic Test	Pass	11:17:24
Air Blank	0.000	11:17:51
0.100 Cal Check	RFI*	11:18:05
Air Blank	IPS**	11:18:18

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

4 of 7 sc
04

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002760
11/14/2011 11:18:26

Standard Lot#: 04911100A1
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:19:09
Diagnostic Test	Pass	11:19:36
Air Blank	0.000	11:20:03
0.100 Cal Check	0.098	11:20:22
Air Blank	0.000	11:20:52
Subject Test	0.097*	11:24:00
Air Blank	0.000	11:24:32
Air Blank	0.000	11:25:00
Subject Test	XXX**	11:25:24
Air Blank	0.000	11:25:58

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002760
11/14/2011 11:26:40

5 of 7 sc

04

Standard Lot#: 04911100A1
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:27:08
Cal Check	0.098	11:27:27
Air Blank	0.000	11:27:56

Sarah Cione

Operator's Signature

EXHIBIT G-2

6 of 7 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

DM
11-30-11

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 11/14/11 TIME 1110
Intoxilyzer Serial # 80-002760 Location DUJHQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)


INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002760
11/14/2011 11:10:40

7 of 7 sc
cu

Standard Lot#: 04911100A1
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:11:08
Cal Check	0.097	11:11:26
Air Blank	0.000	11:11:56



Operator's Signature