CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TSY
9-13-2016

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002760 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 8-25-16 Cal 1 0.098 AC Cal 2 0.097 AC

Between the dates of 7-22-16 and 8-25-16 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Amy L. Autret

9-1-16

ADPS QAS Permit # 33718

SUBSCRIBED AND SWORN to before me this 1st, day of September, 2016.

Elma Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Nov. 3, 2019

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7 TSG 9-13-2016

QA SPECIALIST Amy Autret 53546 AGENCY Tucson Police Department

DATE 8-25-16 TIME 1241

INTOXILYZER SERIAL # 80-002760 LOCATION Crime Laboratory

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS >31 days between checks

SIGNATURE [signature]

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
AA
TB4

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002760
Core Version: 8105.48
08/25/2016 12:44:29

Standard Lot#: 12415100A4
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:45:29
Diagnostic Test	Pass	12:46:04
Air Blank	0.000	12:46:31
0.100 Cal Check	0.098	12:46:52
Air Blank	0.000	12:47:21
Subject Test	0.000	12:47:39
Air Blank	0.000	12:48:06
Wait	ABT*	
Air Blank	0.000	12:48:42

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7
AA
T04

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002760
Core Version: 8105.48
08/25/2016 12:48:46

Standard Lot#: 12415100A4
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:49:38
Diagnostic Test	Pass	12:50:13
Air Blank	0.000	12:50:41
0.100 Cal Check	RFI*	12:50:58
Air Blank	IPS**	12:51:07

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002760
Core Version: 8105.48
08/25/2016 12:51:13

Standard Lot#: 12415100A4
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:52:05
Diagnostic Test	Pass	12:52:41
Air Blank	0.000	12:53:08
0.100 Cal Check	0.098	12:53:29
Air Blank	0.000	12:53:58
Subject Test	0.066*	12:57:05
Air Blank	0.000	12:57:38
Air Blank	0.000	12:58:05
Subject Test	XXX**	12:58:30
Air Blank	0.000	12:59:04

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002760
Core Version: 8105.48
08/25/2016 12:59:29

5 of 7

Standard Lot#: 12415100A4
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:59:58
0.100 Cal Check	0.097	13:00:19
Air Blank	0.000	13:00:48

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
TSH
9.13.2016

QA Specialist Amy Autret 53546 AGENCY Tucson Police Department

DATE 8-25-16 TIME 1241

Intoxilyzer Serial # 80-002760 Location Crime Laboratory

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "~~PUSH BUTTON TO START~~" Contact QAS 31 Day Prev Maintenance"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature]

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002760
Core Version: 8105.48
08/25/2016 12:41:12

Standard Lot#: 12415100A4
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:41:44
0.100 Cal Check	0.098	12:42:05
Air Blank	0.000	12:42:35

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.