CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

TSG 1-16-2008

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002803 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1/11/08 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 12/14/07 and 1/11/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Sarah L. Cione

Sarah L. Cione

1/16/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 16, day of January, 2008.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

1 of 75C

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

154
1-16-2008

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 1/11/08 TIME 10:48am

INTOXILYZER SERIAL # 80-002803 LOCATION Team 2

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
(✓) Deficient sample printed.
( ) 4. Standard Calibration Check standard 0.100 AC
Result: 0.100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Sarah Cione

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7 SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
01/11/2008 10:49:43

Standard Lot#: 730901I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:50:34
Diagnostic Test	Pass	10:51:01
Air Blank	0.000	10:51:28
0.100 Cal Check	0.100	10:51:47
Air Blank	0.000	10:52:17
Subject Test	0.000	10:52:39
Air Blank	0.000	10:53:06
Five Minute Wait	ABT*	
Air Blank	0.000	10:53:47

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7 SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
01/11/2008 10:53:54

Standard Lot#: 730901I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:54:42
Diagnostic Test	Pass	10:55:10
Air Blank	RFI*	10:55:18
Air Blank	IPS**	10:55:26

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1307, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 SC

104

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
01/11/2008 10:55:32

Standard Lot#: 730901I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:56:18
Diagnostic Test	Pass	10:56:46
Air Blank	0.000	10:57:13
0.100 Cal Check	0.101	10:57:32
Air Blank	0.000	10:58:01
Subject Test	0.386*	11:01:10
Air Blank	0.000	11:01:43
Air Blank	0.000	11:02:11
Subject Test	XXX**	11:02:34
Air Blank	0.000	11:03:09

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Pursuant to section 28-1307, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Not a Successfully Completed Test Sequence

5 of 7 SC

154

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
01/11/2008 11:04:07

Standard Lot#: 730901I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:04:34
Cal Check	0.100	11:04:53
Air Blank	0.000	11:05:23

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

6 of 7 SC

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSY 1.16.2008

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 1/11/08 TIME 10:48am

Intoxilyzer Serial # 80-002803 Location Team 2

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7sc

154

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
01/11/2008 10:48:11

Standard Lot#: 730901I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:48:38
Cal Check	0.101	10:48:57
Air Blank	0.000	10:49:27

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD FOR CALIBRATION GAS TANK CHANGE

TSH 1.16.2008

Date: 1/11/2008

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-002803

New Lot Number: 730901 I

New Expiration Date: Nov 6, 2009

(✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 2. Remove the old gas tank and install the new one.

(✓) 3. Hit the "ESC" key twice and swipe the QAS card.

(✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:

Enter Target Value: 0.100
Enter Control Lot #: From the tank label
Enter Exp Date: From the tank label (mm/dd/yyyy)
Edit info Y/N? Y or N
Update Std Info? Y ENTER
"Please wait saving..."

(✓) 5. Hit the "ESC" key twice.

(✓) 6. Go to the "M-Maintenance" menu and select "T-Control Testing." Select "D" for dry control test. After selection is made press ENTER.

(✓) 7. Air blank completed.

(✓) 8. Calibration check completed. Test results 0.101 AC.

(✓) 9. Air blank completed.

(✓) 10. Remove and sign printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 25C

T84

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
01/11/2008 10:46:27

Standard Lot#: 730901I
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:46:55
Cal Check	0.101	10:47:13
Air Blank	0.000	10:47:43

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.