



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSy
1.12.2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002803 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 01/07/10 Cal 1 0.099 AC Cal 2 0.099 AC

Between the dates of 12/10/09 and 01/07/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub

01/11/10

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 11th, day of January, 2010.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

187
CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY
1-12-2010

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 01/07/10 TIME 07:23

INTOXILYZER SERIAL # 80-002803 LOCATION TEAM II

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Cheryl Straub

287
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
01/07/2010

07:25:20

B4

Standard Lot#: 826001I
Last Changed By: T. GALLEGOS #28847

OAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:26:19
Diagnostic Test	Pass	07:26:47
Air Blank	0.000	07:27:14
0.100 Cal Check	0.099	07:27:33
Air Blank	0.000	07:28:03
Subject Test	0.000	07:29:06
Air Blank	ABT*	07:29:13
Air Blank	0.000	07:29:41

*Sequence Aborted

Not a Successfully Completed Test Sequence

397
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
01/07/2010 07:30:07

BM

Standard Lot#: 826001I
Last Changed By: T. GALLEGOS #28847

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SOAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	07:30:49
Air Blank	IPS**	07:30:59

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

427
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
01/07/2010

07:31:20

B4

Standard Lot#: 826001I
Last Changed By: T. GALLEGOS #28847

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:32:11
Diagnostic Test	Pass	07:32:38
Air Blank	0.000	07:33:06
0.100 Cal Check	0.100	07:33:25
Air Blank	0.000	07:33:54
Subject Test	0.118*	07:37:03
Air Blank	0.000	07:37:36
Air Blank	0.000	07:38:03
Subject Test	XXX**	07:39:07
Air Blank	0.000	07:39:41

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

587
CS


INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
01/07/2010 07:40:02

Standard Lot#: 826001I
Last Changed By: T. GALLEGOS #28847

TS4

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	07:40:29
Cal Check	0.099	07:40:48
Air Blank	0.000	07:41:18



Operator's Signature

687
CSTHIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TS4
1-12-2010

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department
 DATE 01/07/10 TIME 07:23
 Intoxilyzer Serial # 80-002803 Location TEAM II

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
 OR
 Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

797
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
01/07/2010 07:23:23

154

Standard Lot#: 826001I
Last Changed By: T. GALLEGOS #28847

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	07:23:50
Cal Check	0.099	07:24:09
Air Blank	0.000	07:24:39



Operator's Signature