CERTIFICATION by CUSTODIAN of RECORDS

*Tucson Police Department Crime Laboratory*
*270 S. Stone Avenue Tucson, AZ 85701-1917*

TSY 10-18-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002803 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10-14-10 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 9-16-10 and 10-14-10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

10-15-10

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 15, day of October, 2010.

Laura B Hodgson

Notary Public

OFFICIAL SEAL
LAURA B. HODGSON
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Aug. 26, 2013

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
CW
TSH
10.18.2010

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 10-14-10 TIME 0752

INTOXILYZER SERIAL # 80-002803 LOCATION ODW

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

( ✓ ) 1. Clock time check.
( ✓ ) 2. Date check.

OPERATIONAL TESTS

( ✓ ) 1. Alcohol-free subject test result 0.000 AC.
( ✓ ) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
( ✓ ) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
( ✓ ) 4. Standard Calibration Check standard 0.100 AC
Result: 0.097 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
10/14/2010 07:54:20

2507
CW
TS4

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 07:55:09 |
| Diagnostic Test | Pass | 07:55:36 |
| Air Blank | 0.000 | 07:56:03 |
| 0.100 Cal Check | 0.098 | 07:56:22 |
| Air Blank | 0.000 | 07:56:52 |
| Subject Test | 0.000 | 07:57:22 |
| Air Blank | ABT* | 07:57:28 |
| Air Blank | 0.000 | 07:57:57 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3567
CM
TS4

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
10/14/2010 07:58:03

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 07:58:50 |
| Diagnostic Test | Pass | 07:59:18 |
| Air Blank | 0.000 | 07:59:45 |
| 0.100 Cal Check | 0.097 | 08:00:04 |
| Air Blank | 0.000 | 08:00:33 |
| Subject Test | RFI* | 08:00:47 |
| Air Blank | IPS** | 08:00:59 |

*RFI Detect
**Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002803
10/14/2010 08:01:04

4 of 7
CM
TSH

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:01:53 |
| Diagnostic Test | Pass | 08:02:21 |
| Air Blank | 0.000 | 08:02:48 |
| 0.100 Cal Check | 0.097 | 08:03:07 |
| Air Blank | 0.000 | 08:03:37 |
| Subject Test | 0.159* | 08:06:45 |
| Air Blank | 0.000 | 08:07:18 |
| Air Blank | 0.000 | 08:07:46 |
| Subject Test | XXX** | 08:08:06 |
| Air Blank | 0.000 | 08:08:41 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
10/14/2010 08:08:57

---

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

---

QAS: C. CONLEY #34517
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:09:24 |
| Cal Check | 0.097 | 08:09:43 |
| Air Blank | 0.000 | 08:10:12 |

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

6567
cu

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSG 10.18.2010

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 10-14-10 TIME 0752

Intoxilyzer Serial # 80-002803 Location ODW

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002803
10/14/2010 07:52:21

7867
lu
TS4

---

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

---

QAS: C. CONLEY #34517
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 07:52:48 |
| Cal Check | 0.097 | 07:53:07 |
| Air Blank | 0.000 | 07:53:37 |

Clare Conley

Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.