



CERTIFICATION by CUSTODIAN of RECORDS  
 Tucson Police Department Crime Laboratory  
 1306 W. Miracle Mile, Tucson, AZ 85705

TS4  
 6-3-2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-006048 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5/20/14 Cal 1 - AC Cal 2 - AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 1/2/14.  
 The instrument was returned to service on \_\_\_\_\_.

Sarah Cione

Sarah L. Cione


5/29/14

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 29<sup>th</sup> day of May, 2014.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

1 of 2  
sc  
TS4  
6-3-2014

Date: 5/20/14

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-006048

Reason For Service: During initial Diagnostics, EEPROM checksum test failed. Will not go to "Ready Mode". Notified of issue by Officer Nolan.

Due to futile attempt to fix another intox w/ same issue, did not attempt to fix this one.

Date Out of Service: 1/2/14

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:  Sent to CMI, Inc.  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_



# Service Evaluation Form

2 of 2 sc  
T34

This form **MUST** be completed and enclosed with instrument to be serviced.  
Failure to complete and return this form may cause delays in service.

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)  
 Name Terry Gallegos Phone: (520) 837-7677  
 Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address: Ship to Address:  
Finance / Accounts Payable Tucson Police Department Crime Lab  
Po Box 27450 1306 W. Miracle Mile  
Tucson, AZ 85726-7450 Tucson, AZ 85705

3. Serial Number: 80-006048 Instrument Model: I8000

4. Detailed Description of Problem:  
During initial diagnostics, EEPROM checksum test failed.

\*\*\*Hazardous Material Warning! - DO NOT return gas cylinders with instrument!\*\*\*

I Authorize Repairs Up To:  All  \$250  \$500  \$750  Other \$ \_\_\_\_\_  
 Purchase Order Number (attach a copy of P.O. if applicable) 34770

Authorized By:

Terry Gallegos Toxicology Coordinator  
 Name (Please Print) Title  
[Signature] 5-20-2014  
 Signature Date

Ship to:  
 CMI, Inc.  
 Attn: Service Dept.  
 316 East Ninth Street  
 Owensboro, KY 42303

\* - Instrument purchased Sept, 2013  
 P.O # 37147  
 No, please send estimate before repairs are made.

Note: An estimate will be faxed before performing any repairs and may cause delays in service.  
 An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.