CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TBJ 5-17-2016

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-006050 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/25/2016 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 3/24/2016 and 4/25/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on ———.
The instrument was returned to service on ———.

William L Loomis

William L. Loomis

5/3/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 3rd, day of May, 2016.

Elma Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Nov. 3, 2019

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 9
WLL
TSK 5.17.2016

QA SPECIALIST W. LOOMIS 53945 AGENCY Tucson Police Department

DATE 4/25/2016 TIME 17:52

INTOXILYZER SERIAL # 80-006050 LOCATION CRIME LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS QAS > 31 days

SIGNATURE Willie Loomis

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 9
WLL

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006050
Core Version: 8105.48
04/25/2016 17:54:20

Standard Lot#: 15615100A1
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: M Weight: 222

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:55:11
Diagnostic Test	Pass	17:55:46
Air Blank	0.000	17:56:13
0.100 Cal Check	0.100	17:56:34
Air Blank	0.000	17:57:04
Subject Test	0.000	17:57:20
Air Blank	0.000	17:57:48
Wait	ABT*	
Air Blank	0.000	17:58:22

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 9 WLL

TBH

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006050
Core Version: 8105.48
04/25/2016 17:58:25

Standard Lot#: 15615100A1
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: M Weight: 222

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:59:13
Diagnostic Test	Pass	17:59:48
Air Blank	0.000	18:00:15
0.100 Cal Check	RFI*	18:00:29
Air Blank	IPS**	18:00:37

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006050
Core Version: 8105.48
04/25/2016 18:00:43

4 of 9 WLL 784

Standard Lot#: 15615100A1
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:01:49
Diagnostic Test	Pass	18:02:24
Air Blank	0.000	18:02:51
0.100 Cal Check	0.102	18:03:12
Air Blank	0.000	18:03:41
Subject Test	0.130*	18:06:49
Air Blank	0.000	18:07:21
Air Blank	0.000	18:07:49
Subject Test	XXX**	18:08:11
Air Blank	0.000	18:08:45

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006050
Core Version: 8105.48
04/25/2016 18:08:59

Standard Lot#: 15615100A1
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	18:09:28
0.100 Cal Check	0.100	18:09:49
Air Blank	0.000	18:10:18

Preventative Maintenance
Performed -- Timer Reset

5 of 9 WLL

TBH

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 9
WLL

TJH
5.17.2016

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department

DATE 4/25/2016 TIME 17:52

Intoxilyzer Serial # 80-006050 Location CRIME LAB

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE William Loomis

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006050
Core Version: 8105.48
04/25/2016 17:52:47

Standard Lot#: 15615100A1
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	17:53:16
0.100 Cal Check	0.101	17:53:37
Air Blank	0.000	17:54:06

Preventative Maintenance
Performed -- Timer Reset

7 of 9
WLL
PH

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

8 of 9
WLL

TSL
5.17.2016

QA Specialist W. LOOMIS AGENCY Tucson Police Department

DATE 4/25/2016 TIME 13:38

Intoxilyzer Serial # 80-006050 Location SALETC

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

CONTACT QAS 31 DAY MAINTENANCE

(✓) 2. Intoxilyzer 8000 display reads "~~PUSH BUTTON TO START~~"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Willie Loomis

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Core Version: 8105.48
Serial Number: 80-006050
04/25/2016 13:38:12

Standard Lot#: 15615100A1
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:38:44
0.100 Cal Check	0.101	13:39:05
Air Blank	0.000	13:39:35

End of test

9 of 9
WLL
TSG

Preventive Maintenance NOT Performed -- Timer NOT Reset

Reprinted From Database
Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.