



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TS4
6.3.2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-006052 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5/20/14 Cal 1 - AC Cal 2 - AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 12/5/13.
The instrument was returned to service on _____.

Sarah Cione

Sarah L. Cione


5/29/14

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 29th, day of May, 2014.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 2sc
TS4
6.3.2014

INTOXILYZER SERVICE RECORD

Date: 5/20/14

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-006052

Reason For Service: During initial Diagnostics, fails EEPROM Checksum Test, Worked with Becky Taylor from CMI on 4/4/14 to fix, but now have Analytical Stability failure.

Date Out of Service: 12/5/13

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:

Sent to CMI, Inc.



Service Evaluation Form

2 of 2 sc
TSH

This form MUST be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)
 Name Terry Gallegos Phone: (520) 837-7677
 Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address: Ship to Address:
Finance / Accounts Payable Tucson Police Department Crime Lab
Po Box 27450 1306 W. Miracle Mile
Tucson, AZ 85726-7450 Tucson, AZ 85705

3. Serial Number: 80-006052 Instrument Model: 8000

4. Detailed Description of Problem:
Originally had Diagnostics fail (EEPROM), worked w/ Becky Taylor on 4/4/14 to fix through modem, but now have Analytical Stability failure.

Hazardous Material Warning! - DO NOT return gas cylinders with instrument!

I Authorize Repairs Up To: All \$250 \$500 \$750 Other \$ _____
 Purchase Order Number (attach a copy of P.O. if applicable) 34770

Authorized By:

Terry Gallegos Toxicology Coordinator
 Name (Please Print) Title

[Signature] 5-20-2014
 Signature Date

★ - Instrument purchased in Sept, 2013
 PO# 37147

No, please send estimate before repairs are made.

Note: An estimate will be faxed before performing any repairs and may cause delays in service.

An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.

Ship to:

CMI, Inc.
 Attn: Service Dept.
 316 East Ninth Street
 Owensboro, KY 42303