



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

*TSG
3.11.2015*

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-006052 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 3-6-15 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of 2-5-15 and 3-6-15 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
 The instrument was returned to service on N/A.


Amy L. Autret
 Amy L. Autret

3-9-15

ADPS QAS Permit # 33718

SUBSCRIBED AND SWORN to before me this 9th, day of March, 2015.

Elma Escarcega
 Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

1 of 7
dla

TJ4
3-11-2015

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Amy Autret 53546 AGENCY Tucson Police Department

DATE 3-6-15 TIME 0934

INTOXILYZER SERIAL # 80-006052 LOCATION ODE - intox room

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.


OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE



2 of 7
dlb
Tsh

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006052
Core Version: 8105.48
03/06/2015 09:38:08

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:39:07
Diagnostic Test	Pass	09:39:44
Air Blank	0.000	09:40:12
0.100 Cal Check	0.100	09:40:32
Air Blank	0.000	09:41:02
Subject Test	0.000	09:41:20
Air Blank	0.000	09:41:47
Wait	ABT*	
Air Blank	0.000	09:42:24

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7
JLB
TSG

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006052
Core Version: 8105.48
03/06/2015 09:43:37

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

DAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:44:35
Diagnostic Test	Pass	09:45:12
Air Blank	0.000	09:45:39
0.100 Cal Check	RFI*	09:45:54
Air Blank	IPS**	09:46:02

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7
AB
734

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006052
Core Version: 8105.48
03/06/2015 09:46:40

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SOAP, SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:47:35
Diagnostic Test	Pass	09:48:12
Air Blank	0.000	09:48:39
0.100 Cal Check	0.100	09:49:00
Air Blank	0.000	09:49:30
Subject Test	0.116*	09:52:37
Air Blank	0.000	09:53:10
Air Blank	0.000	09:53:37
Subject Test	XXX**	09:53:58
Air Blank	0.000	09:54:32

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
dlb
TSL

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006052
Core Version: 8105.48
03/06/2015 09:56:04

Standard Lot#: 25913100A2
Last Changed By: S. CLONE #50616

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:56:34
0.100 Cal Check	0.100	09:56:55
Air Blank	0.000	09:57:25

Preventative Maintenance
Performed -- Timer Reset

leaf 7
TS4
3.11.2015

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Amy Autret 53546 AGENCY Tucson Police Department
DATE 3-6-15 TIME 0934
Intoxilyzer Serial # 80-006052 Location ODE - intox room

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7
JLB

TB4



INTOXILYZER 8000
Instrument Initialization
08:57:05 03/06/2015

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006052
Core Version: 8105.48
03/06/2015 09:34:15

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:34:46
0.100 Cal Check	0.100	09:35:07
Air Blank	0.000	09:35:36

Preventative Maintenance
Performed -- Timer Reset

