



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TS4
4.29.2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-006053 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/24/14 Cal 1 0.099 AC Cal 2 0.097 AC

Between the dates of 3/28/14 and 4/24/14 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


4/28/14

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 28th, day of April, 20 14.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7 sc
TSK
4.29.2014

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 4/24/14 TIME 759

INTOXILYZER SERIAL # 80-006053 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
04/24/2014

08:10:43

2 of 7_{sc}
B4

Standard Lot#: 25913100A2
Last Changed By: A. MARQUIS #47674

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:11:28
Diagnostic Test	Pass	08:12:03
Air Blank	0.000	08:12:30
0.100 Cal Check	0.099	08:12:51
Air Blank	0.000	08:13:21
Subject Test	0.000	08:13:55
Air Blank	0.000	08:14:22
Wait	ABT*	
Air Blank	0.000	08:15:36

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
04/24/2014

08:18:14

3 of 7
sc
TBH

Standard Lot#: 25913100A2
Last Changed By: A. MARQUIS #47674

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:19:10
Diagnostic Test	Pass	08:19:44
Air Blank	0.000	08:20:12
0.100 Cal Check	0.097	08:20:32
Air Blank	0.000	08:21:02
Subject Test	RFI*	08:21:20
Air Blank	IPS**	08:21:30

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
04/24/2014

08:25:04

4 of 7
SC
TS4

Standard Lot#: 25913100A2
Last Changed By: A. MARQUIS #47674

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:25:57
Diagnostic Test	Pass	08:26:32
Air Blank	0.000	08:26:59
0.100 Cal Check	0.098	08:27:20
Air Blank	0.000	08:27:50
Subject Test	0.155*	08:30:57
Air Blank	0.000	08:31:29
Air Blank	0.000	08:31:57
Subject Test	XXX**	08:32:30
Air Blank	0.000	08:33:04

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
04/24/2014 08:33:28

5 of 7
sc
TSy

Standard Lot#: 25913100A2
Last Changed By: A. MARQUIS #47674

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:33:57
0.100 Cal Check	0.097	08:34:18
Air Blank	0.000	08:34:47

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6077
134 SC
4.29.2014

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 4/24/14 TIME 759
Intoxilyzer Serial # 80-006053 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
04/24/2014 07:59:21

7 of 7 SC
BY

Standard Lot#: 25913100A2
Last Changed By: A. MARQUIS #47674

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	07:59:49
0.100 Cal Check	0.099	08:00:10
Air Blank	0.000	08:00:40

Preventative Maintenance
Performed -- Timer Reset