



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

AA
 9-19-16

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-006053 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 9-8-2016 Cal 1 0.098 AC Cal 2 0.097 AC

Between the dates of — and — the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —
 The instrument was returned to service on 9-8-2016

Terry Gallegos

Terry S. Gallegos


9-15-2016

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 15th, day of September, 2016.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

Handwritten: 9-19-16

INTOXILYZER SERVICE RECORD

Handwritten: 10/12
TJ

QAS: Terry Gallegos 28847

Intoxilyzer Serial Number: 80-006053

Date Returned to Service: 9-8-2016

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

MB
9-19-16
Zof12
TS4

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 9-8-2016 TIME 1655

INTOXILYZER SERIAL # 80-006053 LOCATION CrimeLab

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE *Terry Gallegos*

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
09/08/2016

16:57:28

3 of 12
TS4
db

Standard Lot#: 06616100A1
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:58:39
Diagnostic Test	Pass	16:59:14
Air Blank	0.000	16:59:42
0.100 Cal Check	0.098	17:00:03
Air Blank	0.000	17:00:32
Subject Test	0.000	17:00:52
Air Blank	0.000	17:01:19
Wait	ABT*	
Air Blank	0.000	17:01:53

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
09/08/2016 17:02:00

4 of 12
TS4
[Signature]

Standard Lot#: 06616100A1
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:02:47
Diagnostic Test	Pass	17:03:22
Air Blank	0.000	17:03:49
0.100 Cal Check	RFI*	17:04:04
Air Blank	IPS**	17:04:11

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
09/08/2016

17:04:34

50f12
Tsh
[Signature]

Standard Lot#: 06616100A1
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:05:19
Diagnostic Test	Pass	17:05:54
Air Blank	0.000	17:06:21
0.100 Cal Check	0.097	17:06:42
Air Blank	0.000	17:07:12
Subject Test	0.118*	17:10:19
Air Blank	0.000	17:10:52
Air Blank	0.000	17:11:20
Subject Test	XXX**	17:11:32
Air Blank	0.000	17:12:06

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
09/08/2016 17:12:18

6 of 12
TS4
[Signature]

Standard Lot#: 06616100A1
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	17:12:47
0.100 Cal Check	0.097	17:13:00
Air Blank	0.000	17:13:38
Air Blank	0.000	17:16:06

*Input Timed Out

EXHIBIT G-2


THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

7 of 12
TS4
9-19-16

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 9-8-2016 TIME 1655
Intoxilyzer Serial # 80-006053 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-006053
Core Version: 8105.48
09/08/2016 16:55:42

80f12
TS4
[Signature]

Standard Lot#: 06616100A1
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:56:14
0.100 Cal Check	0.098	16:56:34
Air Blank	0.000	16:57:04

Preventative Maintenance
Performed -- Timer Reset



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

9 of 12
TS4
9-19-16

Date: 9-8-2016
 QAS: Terry Gallegos 28847
 Intoxilyzer Serial Number: 80-006053
 New Lot Number: 06616100A1
 New Expiration Date: 4-5-2018

1. Intoxilyzer 8000 display reads "Contact QAS std change required" ~~"PUSH BUTTON TO START"~~
2. Remove the old gas tank and install the new one.
3. Hit the "ESC" key twice and swipe the QAS card.
4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:
 Enter Target Value: 0.100
 Enter Control Lot #: From the tank label
 Enter Exp Date: From the tank label (mm/dd/yyyy)
 Edit info Y/N? Y or N
 Update Std Info? Y ENTER
 "Please wait saving..."
5. Hit the "ESC" key twice.
6. Go to the "M-Maintenance" menu and select "T-Control Testing." Select "D" for dry control test. After selection is made press ENTER.
7. Air blank completed.
8. Calibration check completed. Test results 0. 098 AC.
9. Air blank completed.
10. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 90 006052
Core Version: 8105.48
09/08/2016 16:49:27

10 of 12
TSL
[Signature]

Standard Lot#: 06616100A1
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:50:00
0.100 Cal Check	0.098	16:50:21
Air Blank	0.000	16:50:50

Preventative Maintenance
Performed -- Timer Reset

120f1234

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-006053, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58 No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date August 16, 2006 Signed [Signature]



Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev.A