7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 3087

Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 09312100A1
Expiration: 6/1/2014

0.100 BAC (For use with breath alcohol testing instruments)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	260.5 ppm	+/- 0.002 or 2% BAC whichever is greater	NDIR
Nitrogen	Balance		

*NIST Standard Reference Material
Certification of NTRM Batch No. 091602
Nominal 210 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Specialty Gas Lab Tech

5-2-12
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

ISO/IEC 17025:2005 Accredited Laboratory