



ISO/IEC 17025:2005 Accredited Laboratory

Certificate of Analysis

Certificate ID: 8097
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 12415100A4
Expiration: 6/5/2017


0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	260 ppm	+/- 0.002 or 2% BAC whichever is greater	NDIR
Nitrogen	Balance		

*NIST Standard Reference Material
 Cylinder No. CCI4290 / Job No. 09160202
 Certified 212.8 µmol/mol Ethanol in Nitrogen
 for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition
 and direct sunlight. Do not allow storage area to
 exceed 52 °C (125 °F).


 Specialty Gas Lab Tech

05/20/15
 Date



Distributed by:

CMI Inc.
 316 East Ninth Street
 Owensboro, KY 42303
 Phone 866-835-0690
www.alcoholtest.com