ILMO specialty gases

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

# Certificate of Analysis

**Certificate ID:** 5297

**Part #:** BAC105L100T
**Cylinder Size:** 105L
**Lot Number:** 19613100A4
**Expiration:** 8/1/2015

## 0.100 BAC (For use with breath alcohol testing instruments)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

| Component: | Concentration: | Accuracy: | Method: |
|---|---|---|---|
| Ethanol | 260 ppm | +/- 0.002 or 2% BAC whichever is greater | NDIR |
| Nitrogen | Balance | | |

*NIST Standard Reference Material
Cylinder No. CC14290 / Job No. 09160202
Certified 212.8 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Jacob Matter
Specialty Gas Lab Tech

07/30/13
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

ISO/IEC 17025:2005 Accredited Laboratory