Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS
EBS™ - Ethanol Breath Standard

Customer Name	Assay Laboratory	Purchase Order : 367651
CMI	Scott Specialty Gases	Scott Project #: 0119378
316 East Ninth Street	6141 Easton Road, Bldg. 1	Lot No.: 4323021
Owensboro, KY 42303	Plumsteadville, Pa.	Expiration Date:
	18949-0310	November 23, 2006

REFERENCE STANDARD	CYLINDER #	CONCENTRATION ppm
N.I.S.T. Traceable Standards*	ALM-060154	103.6
	ALM-060204	118.1
	ALM-059962	222.2
	ALM-037472	251.0

CONTAINER TYPE: 105

LOT CERTIFIED VALUE

CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	260.5 (0.100)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst: Floyd Bard

Date: 11/23/2004

b:coa1118.doc