Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS
EBS™ - Ethanol Breath Standard

Customer Name	Assay Laboratory	Purchase Order : 368147
CMI	Scott Specialty Gases	Scott Project #: 0126184
316 East Ninth Street	6141 Easton Road, Bldg. 1	Lot No.: 507501G
Owensboro, KY 42303	Plumsteadville, Pa.	Expiration Date:
	18949-0310	March 18, 2007

REFERENCE STANDARD	CYLINDER #	CONCENTRATION ppm
N.I.S.T. Traceable Standards*	ALM-060154	103.6
	ALM-060204	118.1
	ALM-059962	222.2
	ALM-037472	251.0

CONTAINER TYPE: 105

LOT CERTIFIED VALUE
CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	260.5 (0.100)	+- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst: Floyd Boof

Date: 03/18/2005

b:coa1118.doc