Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS
EBS™ - Ethanol Breath Standard

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order: 370535
Scott Project # 0159921
Lot No.: 629202I
Expiration Date:
October 23, 2008

REFERENCE STANDARD	CYLINDER #	CONCENTRATION ppm
N.I.S.T. Traceable Standards*	ALM-059986	103.2
	ALM-060143	118.1
	ALM-060196	223.2
	ALM-037475	302.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE
CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	260.5 (0.100)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Floyd Bard

Analyst:

Date: 10/23/2006

b:coa1118.doc