Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS
EBS™ - Ethanol Breath Standard

Customer Name	Assay Laboratory	Purchase Order: 370657
CMI	Scott Specialty Gases	Scott Project # 0164250
316 East Ninth Street	6141 Easton Road, Bldg. 1	Lot No.: 630601I
Owensboro, KY 42303	Plumsteadville, Pa. 18949	Expiration Date: November 6, 2008

REFERENCE STANDARD	CYLINDER #	EXPIRATION DATE	CONCENTRATION PPM
N.I.S.T. Traceable Standards*	ALM-049986	05/18/2007	262.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE
CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	260.5 (0.100)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Floyd Bard

Analyst: Date: 11/27/2006

b:coa1118.doc