Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS
EBS™ - Ethanol Breath Standard

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order:
370720/REPLACEMENT
Scott Project # 0187984
Lot No.: 635301I
Expiration Date:
December 21, 2008

REFERENCE STANDARD	CYLINDER #	EXPIRATION DATE	CONCENTRATION PPM
N.I.S.T. Traceable Standards*	ALM-049986	05/18/2007	262.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE
CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	260.5 (0.100)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Floyd Bard

Analyst:

Date: 02/05/2007

b:coa1118.doc