



CERTIFICATE OF ANALYSIS
EBS - ETHANOL BREATH STANDARD

CMI
316 EAST NINTH STREET
OWENSBORO, KY 42303

INVOICE#: 34083372
PO#: 373535
CMI ITEM #: 340135
DATE: 23 Jul 2009

ANALYTICAL ACCURACY: +/- 2%
PRODUCT EXPIRATION: 23 Jul 2011

CALGAZ LOT#: 642845
ETHANOL IN NITROGEN

COMPONENT	PPM	(BAC)
ETHANOL	260.5	(0.100)
NITROGEN	BAL	

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
NIST TRACEABLE STANDARDS*	D752020	261.4

* CERTIFICATION TRACEABLE TO N.I.S.T. RGM ETHANOL STANDARDS

MANUFACTURED DATE: 23 Jul 2009

CALGAZ CYLINDER SIZE: 6D

APPROVED BY :

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, DIV. OF AL ADVANCED TECHNOLOGIES U.S. LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410)228-6400 Fax: (410)228-4251