Scott Specialty Gases

Shipped From:
6141 EASTON ROAD, BLDG 1 PO BOX 310
PLUMSTEADVILLE PA 18949-0310
Phone: 800-331-4953 Fax: 215-766-7226

C E R T I F I C A T E O F A N A L Y S I S
EBS - ETHANOL BREATH STANDARD

CMI

316 EAST NINTH STREET
RECEIVING DEPARTMENT
OWENSBORO KY 42303

PROJECT #: 01-07783-002
PO#: 371008
ITEM #: 0104EBS1000 105
DATE: 12Mar2007

ANALYTICAL ACCURACY: +/- 2%
PRODUCT EXPIRATION: 11Mar2009

SCOTT LOT#: 706602I
ETHANOL IN NITROGEN

COMPONENT	PPM	(BAC) LES)
ETHANOL	260.5	(0.100)
NITROGEN	BAL	

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
NIST TRACEABLE STANDARDS *	ALM049986	262.0

* CERTIFICATION TRACEABLE TO N.I.S.T. RGM ETHANOL STANDARDS

MANUFACTURED DATE: 07Mar2007 SCOTTY SIZE: 105

ANALYST: _____
TOM BARONE