EXHIBIT Q

1086

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran / Missie
(Print Name)

INTOXILYZER SERIAL #(7)04 001225 LOCATION TRAFFIC

DATE Jan. 26 ~~19~~ 2004 TIME 1102

Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .362.

Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

Alcohol free subject Test result 0.000 AC

Error recognition logic system functioning
Invalid test printed

Proper sample recognition system
Invalid sample printed
Deficient sample printed

N/A Completeness of sample purge with collection tube**

Calibration standard 0.080 AC
Results: 0.079AC, 0.080AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Irene Doran Rhonda Missie

0. alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 64-001225 #7 01/26/04
E47.02 11:07
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

TRAFFIC
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225 #7
01/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:02
SUBJECT TEST	.000	11:02
AIR BLANK	.000	11:03

DVM=.362

DORAN
SUBJECT'S NAME

TRAFFIC
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

p. 296

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
01/26/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:12
CAL. CHECK	.079	11:12
AIR BLANK	.000	11:13

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: TRAFFIC

OPERATOR: DORAN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
01/26/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:14
CAL. CHECK	.080	11:14
AIR BLANK	.000	11:14

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: TRAFFIC

OPERATOR: DORAN (46331)

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

p.486

SN 64-001225 #7 01/26/04
E47.02 11:18
INVALID TEST

SUBJECT'S NAME

TRAFFIC

TIME FIRST OBSERVED INSTRUMENT LOCATION

DORAN 46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

pushed button at wrong time

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 1/26/04 Time 1102
Intoxilyzer Serial# (7) 64-001225 Location TRAFFIC
QA Specialist Irene Doran / R Misse
(Print Name)

(JD) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(JD) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(JD) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(JD) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

(JD) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

(JD) 6. Push Start Test button or press enter on keyboard.

(JD) 7. Insert card in response to display.

(JD) 8. Air blank completed.

(JD) 9. Calibration check completed. / 0.080 AC
Test results 0.079 AC

(JD) 10. Air blank completed.

(JD) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(JD) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran Rhonda Misse

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
01/26/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:19
*SUBJECT TEST	.073	11:21
AIR BLANK	.000	11:21

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TRAFFIC
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN 46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001225
01/26/04 #7

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:10
INVALID SAMPLE	.XXX	11:10
AIR BLANK	.000	11:11

DORAN
SUBJECT'S NAME

TRAFFIC
INSTRUMENT LOCATION

TIME FIRST OBSERVED

DORAN (46331)
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

-Binaca-

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P 686