182
15

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

Date: 3/1/04
Intoxilyzer Serial Number: ~~06~~ 64 - 001335 (#7)
QAS: Irene Doran 46331

Problem Description:
Instrument did not adjust date for leap year.

If the instrument is taken out of service, complete the following:

Date out of service N/A

Calibration Check completed prior to removal from service:

☐ Yes Date:_____ ☐ No

SQAP completed prior to removal from service:

☐ Yes Date:_____ ☐ No

Date returned to service: N/A

Corrective Action Taken or Repairs:
Adjusted date from 3/2/04 to 3/1/04 before sending out for repair.

282
45

SN 64-001227
F47.02
INVALID TEST

03/02/04
17:44

#7

Date incorrect for leap yr.

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

#7

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.