



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

184

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 5000 Serial Number 64-001225 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 1/20/05 Cal 1 0.073 AC Cal 2 0.073 AC  
Date: 2/8/05 Cal 1 N/A AC Cal 2 N/A AC  
Date: N/A Cal 1          AC Cal 2          AC

Between the dates of N/A and N/A the instrument was working properly and accurately.


The instrument was taken out of service for repair on 2/8/05.  
The instrument was returned to service on 3/18/05.

Jennifer L. Coombs  
Jennifer L. Coombs

3/23/05

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 23, day of March, 2005.



Kathy A. Kinney  
Notary Public

\*Includes documents specified in A.A.C. R9-14-404.A.6



# MEMORANDUM

**DATE:** February 9, 2005

**TO:** Whom it may concern

**FROM:** Jennifer Coombs  
Intoxilyzer 5000 QAS  
Tucson Police Department  
Crime Lab  
270 S. Stone Avenue  
Tucson, AZ 85701-1917  
(520) 791-4494



**SUBJECT:** Intoxilyzer Out of Service

Intoxilyzer 64-001225 (#7), was taken out of service for repair on **February 8, 2005**.

I cannot testify that it was working properly and accurately between the dates of **January 20, 2005** and **February 8, 2005**.

This instrument was located at **TRAFFIC**.



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

Date: February 9, 2005

QAS: Jennifer Coombs 46660

Intoxilyzer Serial Number: 64-001225 (#7)

Reason For Service: On 2/8/05, Linda from Traffic called to report the instrument displayed "Processor Error 1".

Date Out of Service: 2/8/05 pulled by TSG

Calibration Check completed prior to removal from service

Date: No

SQAP completed prior to removal from service

Date: No

Corrective Action:

Sent to Applied Electronics