182

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

Date: 11/17/04
Intoxilyzer Serial Number: 64-00226 (#8)
QAS: Irene Doran #46331

Problem Description:
Instrument error : Unable to obtain a stable reference reported by Ofc. Flores @ 0300 11/17/04 Instrument located at ODM/C13

If the instrument is taken out of service, complete the following:

Date out of service 11/17/04

Calibration Check completed prior to removal from service:

☐ Yes Date: _____ ☑ No

SQAP completed prior to removal from service:

☐ Yes Date: _____ ☑ No

Date returned to service: 12/13/04

Corrective Action Taken or Repairs: Error reproduced in lab. Sending to Applied Electronics.

SN 64-001226 11/17/04
E47.02 02:57
INVALID TEST
UNABLE TO OBTAIN A
STABLE REFERENCE.

#8

SUBJECT'S NAME

FIRST OBSERVED INSTRUMENT LOC

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

SN 64-001226 11/17/04
E47.02 17:15
INVALID TEST
UNABLE TO OBTAIN A
STABLE REFERENCE.

#8

SUBJECT'S NAME

LAB

TIME FIRST OBSERVED INSTRUMENT LOCATION

46331

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

I picked up instrument 1017 on 11/17/04. [illegible] 46331

292

MEMORANDUM

DATE: 11/18/04

TO: Whom it may concern

FROM: Irene Doran
Intoxilyzer 5000 QAS
Tucson Police Department
Crime Lab
270 S. Stone Avenue
Tucson, AZ 85701-1917
(520) 791-4494

SUBJECT: Intoxilyzer Out of Service

Intoxilyzer **64-001226 (#8)** was taken out of service for repair on **November 17, 2004** .
I cannot testify that it was working properly and accurately between the dates of

10/28/2004 and **11/17/2004** .