CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 64- 001226(8) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 3/24/05 Cal 1 0.079 AC Cal 2 0.079 AC

Date: 4/20/05 Cal 1 0.079 AC Cal 2 0.081 AC

Date: NA Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 3/24/05 and 4/20/05 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on NA.
The instrument was returned to service on NA.

Irene Doran

Irene Doran

4/20/05

ADPS QAS Permit # QAS 29586

SUBSCRIBED AND SWORN to before me this 20, day of April, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

186

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Irene Doran 46331 (Print Name)

INTOXILYZER SERIAL # 64-001226 (#8) LOCATION DUI HQ

DATE April 20 2005 TIME 1058

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .478.

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓✓ Calibration standard 0.080 AC
Results: 0.079AC + 0.081 AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: ___

SIGNATURE Irene Doran

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 64-001226 #8 04/20/05
E47.02 11:15
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 I----------
Q300 I----------
Q305 I----------
Q302 I----------
Q303 I----------
Q304 I----------
Q301 I----------

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
04/20/05

#8

TEST	BAC VALUE	TIME
AIR BLANK	.000	10:59
SUBJECT TEST	.000	10:59
AIR BLANK	.000	11:00

DVM=.478

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
04/20/05

#8

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:09
CAL. CHECK	.079	11:09
AIR BLANK	.000	11:09

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
04/20/05

#8

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:10
CAL. CHECK	.081	11:10
AIR BLANK	.000	11:11

SUBJECT'S NAME

TIME FIRST OBSERVED — INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46331

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

396

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

486
05

SN 64-001226 04/20/03
E47.02 11:01
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

#8

SUBJECT'S NAME

DUI HQ
TIME FIRST OBSERVED INSTRUMENT LOCATION

4633
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

596

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 4/20/05 Time 1058
Intoxilyzer Serial # 64-001226 (#8) Location DUI HQ
QA Specialist Irene Doran 46331
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
(✓) 6. Push Start Test button or press enter on keyboard.
(✓) 7. Insert card in response to display.
(✓) 8. Air blank completed.
(✓) 9. Calibration check completed.
Test results 0.079 + 0.081 AC
(✓) 10. Air blank completed.
(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Irene Doran

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
04/20/05

#8

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:01
INVALID SAMPLE	.XXX	11:02
AIR BLANK	.000	11:02

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

sweet breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
04/20/05

#8

TEST	BAC VALUE	TIME
AIR BLANK	.000	11:12
*SUBJECT TEST	.147	11:14
AIR BLANK	.000	11:14

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46331
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.