CERTIFICATION by CUSTODIAN of RECORDS.

*Tucson Police Department Crime Laboratory*

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 64-001226(#8) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 10/6/05 Cal 1 0.076 AC Cal 2 0.076 AC

Date: 11/2/05 Cal 1 0.077 AC Cal 2 0.077 AC

Date: N/A Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 10/6/05 and 11/2/05 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Jennifer L. Coombs

Jennifer L. Coombs

11/8/05

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 8, day of November, 2005.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1 of 6 jle

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Jennifer Coombs 461160
(Print Name)

INTOXILYZER SERIAL # 64-001226 (#8) LOCATION DUI HQ

DATE November 2 ~~19~~ jle 2005 TIME 1020

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .456 .

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓✓ Calibration standard 0.080 AC
Results: 0.077AC & 0.077AC

Instrument operating properly and accurately. YES ✓ NO ___

COMMENTS: _____

SIGNATURE Jennifer L. Coombs

0. N/A alcohol concentration standard collected for subsequent analysis.**

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**
****THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.**

DHS/DSLS/Form C117 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

SN 64-001226 11/02/05 #8
E47.02 10:27
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC #8
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
11/02/05

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:29 |
| SUBJECT TEST | .000 | 10:29 |
| AIR BLANK | .000 | 10:30 |

DVM=.456

SUBJECT'S NAME

DUI HQ

TIME FIRST OBSERVED INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

2 of 6 file

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

# Tucson Police Department Crime Laboratory

FEDERAL SIGNAL CORP. / CMI INC #8
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
11/02/05

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:31 |
| CAL. CHECK | .077 | 10:31 |
| AIR BLANK | .000 | 10:32 |

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46160

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC #8
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
11/02/05

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:32 |
| CAL. CHECK | .077 | 10:33 |
| AIR BLANK | .000 | 10:33 |

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION: DUI HQ

OPERATOR: 46160

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 6

4 of 6
file

SN 64-001226 11/02/05 #8
E47.02 10:30
INVALID TEST

SUBJECT'S NAME

TIME FIRST OBSERVED

DUI HQ
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

pushed start test button at wrong time

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

NM

5 of 6
file

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 11/2/05 Time 1620
Intoxilyzer Serial # 64-001226 (#8) Location DUI HQ
QA Specialist Jennifer Coombs 46160
(Print Name)

( ✓ ) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC
( ✓ ) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.
( ✓ ) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.
( ✓ ) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.
✓ ( ✓ ) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".
✓ ( ✓ ) 6. Push Start Test button or press enter on keyboard.
✓ ( ✓ ) 7. Insert card in response to display.
✓ ( ✓ ) 8. Air blank completed.
✓ ( ✓ ) 9. Calibration check completed.
Test results 0.077AC & 0.077 AC
✓ ( ✓ ) 10. Air blank completed.
✓ ( ✓ ) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.
( ✓ ) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

***WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD**

DHS/DSLS/Form C116 (Rev. 7-93)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

FEDERAL SIGNAL CORP. / CMI INC #8
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
11/02/05

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:34 |
| INVALID SAMPLE | .XXX | 10:34 |
| AIR BLANK | .000 | 10:35 |

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION DUI HQ

OPERATOR 46160

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC #8
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-001226
11/02/05

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 10:37 |
| *SUBJECT TEST | .142 | 10:39 |
| AIR BLANK | .000 | 10:39 |

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION DUI HQ

OPERATOR 46160

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 6