P. 1of6
jlc

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

AGENCY Tucson Police QA SPECIALIST Jennifer Coombs 461160
(Print Name)

INTOXILYZER SERIAL #64-002505 (#11) LOCATION ODM
DATE July 9 ~~18~~ 2004 jlc TIME 0917

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .271.

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓ Calibration standard 0.080 AC
Results: 0.080 & 0.085AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Jennifer L. Coombs

0.N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

#11

SN 64-002505 07/09/04
E47.02 09:23
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 I----------
Q306 I----------
Q305 I----------
Q302 I----------
Q303 I----------
Q304 I----------
Q301 I----------

SUBJECT'S NAME

ODM

TIME FIRST OBSERVED INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

#11

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
07/09/04

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:25 |
| SUBJECT TEST | .000 | 09:25 |
| AIR BLANK | .000 | 09:25 |

DVM=.271

SUBJECT'S NAME

ODM

TIME FIRST OBSERVED INSTRUMENT LOCATION

461160

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

P. 2 of 6

#11

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
07/09/04

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:20 |
| CAL. CHECK | .080 | 09:20 |
| AIR BLANK | .000 | 09:21 |

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

#11

FEDERAL SIGNAL CORP. / CMI INC
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
07/09/04

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:21 |
| CAL. CHECK | .085 | 09:22 |
| AIR BLANK | .000 | 09:22 |

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

P. 396
file

P. 4 of 6
file

SN 64-002505
E47.02
INVALID TEST

07/09/04 #11
09:26

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Pushed start test button at wrong time

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 7/9/04 Time 0917
Intoxilyzer Serial# 64-002505(#11) Location ODM
QA Specialist Jennifer Coombs 461160
(Print Name)

( ✓ ) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

( ✓ ) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

( ✓ ) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

( ✓ ) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

✓ ( ✓ ) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

✓ ( ✓ ) 6. Push Start Test button or press enter on keyboard.

✓ ( ✓ ) 7. Insert card in response to display.

✓ ( ✓ ) 8. Air blank completed.

✓ ( ✓ ) 9. Calibration check completed.
Test results 0.080 & 0.085 AC

✓ ( ✓ ) 10. Air blank completed.

✓ ( ✓ ) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

( ✓ ) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
07/09/04

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:27 |
| INVALID SAMPLE | .XXX | 09:27 |
| AIR BLANK | .000 | 09:28 |

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
07/09/04

| TEST | BAC VALUE | TIME |
|---|---|---|
| AIR BLANK | .000 | 09:29 |
| *SUBJECT TEST | .062 | 09:31 |
| AIR BLANK | .000 | 09:31 |

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

461160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

P. 6 of 6
fc