1 of 6
file

EXHIBIT Q

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000* 184

AGENCY Tucson Police QA SPECIALIST Jennifer Coombs 461160
(Print Name)

INTOXILYZER SERIAL # 64-002505 (#11) LOCATION ODM

DATE November 26 ~~29~~ 2004 TIME 0846

✓ Display reads "READY TO START" or "PUSH BUTTON TO START TEST."

DIAGNOSTIC TESTS

✓ DVM Test check. Setting should be between .200 and .600. Mode selection switch S2 on, S1 and S3 off or keyboard menu selection "H". Reading is .307.

✓ Display Test check. Mode selection switch S1 on and S2 and S3 off or keyboard menu selection "V".

✓ Printer Test check. Mode selection switch S1, S2, S3 off or keyboard menu selection "P".

✓ Clock time check. Mode selection switch S10 on or keyboard menu selection "E".

✓ Date check. Mode selection switch S11 on or keyboard menu selection "E".

OPERATIONAL TESTS

✓ Alcohol free subject Test result 0.000 AC

✓ Error recognition logic system functioning
Invalid test printed

✓ Proper sample recognition system
Invalid sample printed ✓
Deficient sample printed ✓

N/A Completeness of sample purge with collection tube**

✓ Calibration standard 0.080 AC
Results: 0.074AC & 0.074AC

Instrument operating properly and accurately. YES ✓ NO__

COMMENTS: _____

SIGNATURE Jennifer L. Coombs

0. N/A alcohol concentration standard collected for subsequent analysis.**

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD
**THIS STEP IS ONLY REQUIRED IF THIS DEVICE IS BEING USED FOR SAMPLE CAPTURE.

DHS/DSLS/Form C117(Rev.7-93)

SN 64-002505 11/26/04 #11
E47.02 08:56

EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
EEEEEEEEEEEEEEEEEEEEEEEEEEEEEEEE
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
HHHHHHHHHHHHHHHHHHHHHHHHHHHHHHHH
Q307 |----------
Q306 |----------
Q305 |----------
Q302 |----------
Q303 |----------
Q304 |----------
Q301 |----------

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
11/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	08:54
SUBJECT TEST	.000	08:55
AIR BLANK	.000	08:55

DVM = .307

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

© 1986 by CMI INC.

CMI INC

2 of 6

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
11/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	08:50
CAL. CHECK	.074	08:50
AIR BLANK	.000	08:51

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
11/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	08:51
CAL. CHECK	.074	08:52
AIR BLANK	.000	08:52

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

Tucson Police Department Crime Laboratory

3 of 6
file

4 of 6
file

SN 64-002505
E47.02
INVALID TEST
SAMPLE INTRODUCED
AT IMPROPER TIME.

11/26/04 #71
08:57

SUBJECT'S NAME

TIME FIRST OBSERVED

ODM
INSTRUMENT LOCATION

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC.

© 1986 by CMI INC.

5 of 6
file

EXHIBIT P

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R9-14-404(A)

ARIZONA DEPARTMENT OF HEALTH SERVICES

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 5000*

STANDARD CALIBRATION CHECK PROCEDURE

Agency Tucson Police Department Date 11/26/04 Time 0846
Intoxilyzer Serial# 64-002505 (#11) Location ODM
QA Specialist Jennifer Coombs 461160
(Print Name)

(✓) 1. Pour a standard alcohol solution of known value into a clean dry simulator jar and assemble the simulator. Insure that a tight seal has been made.
Standard value: 0.080 AC

(✓) 2. Turn on the simulator and allow the temperature to reach 34°C±.2°C.

(✓) 3. Set Intoxilyzer mode selection in the ACA mode by switching mode selection switch #9 on or selecting "C" on keyboard menu.

(✓) 4. Attach simulator to the simulator entrance port on the Intoxilyzer.

✓(✓) 5. Intoxilyzer 5000 display reads "READY TO START" or "PUSH BUTTON".

✓(✓) 6. Push Start Test button or press enter on keyboard.

✓(✓) 7. Insert card in response to display.

✓(✓) 8. Air blank completed.

✓(✓) 9. Calibration check completed.
Test results 0.074AC ±0.07H AC

✓(✓) 10. Air blank completed.

✓(✓) 11. When display reads Test Complete remove evidence card. Attach the card to the completed checklist.

(✓) 12. Return mode selection switch #9 to off position after all calibration checks are complete or type Q and enter on keyboard.

SIGNATURE Jennifer L. Coombs

*WITH OR WITHOUT VAPOR RECIRCULATION AND WITH OR WITHOUT KEYBOARD

DHS/DSLS/Form C116 (Rev.7-93)

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
11/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	08:58
INVALID SAMPLE	.XXX	08:58
AIR BLANK	.000	08:58

SUBJECT'S NAME

ODM
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

Sweet Breath

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

FEDERAL SIGNAL CORP. / CMI INC #11
INTOXILYZER - ALCOHOL ANALYZER
AZ MODEL 5000 SN 64-002505
11/26/04

TEST	BAC VALUE	TIME
AIR BLANK	.000	08:59
*SUBJECT TEST	.081	09:01
AIR BLANK	.000	09:01

* DEFICIENT SAMPLE - VALUE PRINTED WAS HIGHEST OBTAINED.

SUBJECT'S NAME

ODM
INSTRUMENT LOCATION

TIME FIRST OBSERVED

46160
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD

CMI INC

© 1986 by CMI INC.

6 of 6