



TUCSON POLICE DEPARTMENT
CRIME LABORATORY
INTOXILYZER SERVICE RECORD

Date: 01.01.2004
Intoxilyzer Serial Number: #12 64-002516
QAS: Terry Gallegas 28847

Problem Description:
Change date to 2004

If the instrument is taken out of service, complete the following:

Date out of service _____

Calibration Check completed prior to removal from service:

Yes Date: _____ No

SQAP completed prior to removal from service:

Yes Date: _____ No

Date returned to service: _____

Corrective Action Taken or Repairs:
Open side panel and rotated year dial to "04."